

Matrix optimization for the MALDI-TOF-MS analysis of trace biodiesel components

Casey R. McAlpin,^{1,2} Kent J. Voorhees,¹ Teresa L. Alleman,² and Robert L. McCormick²

NREL
National Renewable
Energy Laboratory
Innovation for Our Energy Future

Department of Chemistry, Colorado School of Mines, 1500 Illinois Street, Golden, CO 80401 • Center for Transportation Technologies and Systems, National Renewable Energy Laboratory, 1617 Cole Blvd., Golden, CO 80401

Overview

- Cold weather operability is a limiting factor in expanding the use of biodiesel. At low temperatures, biodiesel can form precipitates, thought to be complex mixtures of trace components.
- Previously, trace components were problematic to characterize because of the complex nature of the fuel.
- The optimized MALDI-TOF method was designed to circumvent problems encountered with traditional analyses.
- The optimized matrix system was hypothesized to be a "catch-all" matrix for ionizing the range of trace compounds in biodiesel.
- Standards and mixtures of standards were analyzed to represent the range of compounds present in biodiesel.
- Optimized MALDI-TOF analysis can be used as a highthroughput fuel screening tool.
- TOF/TOF mass analysis can be performed to verify peak assignment of compounds present in mixtures and neat fuels.





Figure 1. The spectra above are of monostearin (A) and distearin (B) analyzed with Nal-doped HCCA and Nal-doped optimized matrix mixture (front and back, respectively). The sodiated ions are absent from the spectra obtained with only HCCA and Nal.






Figure 3. The spectrum (above left) monostearin selected from the monoacylglyceride "region" of the spectrum. The monostearin was then extracted for CID analysis, resulting in the observed fragmentation spectrum (bottom left). The peaks labeled A-K are identified in the table at right.

Experimental

- Standards of mono-, di-, and triacylglycerides were obtained, as well as standards of free sterols
 and sterol glucosides representative of compounds present as trace components in biodiesel.
- Samples of standards were analyzed first with traditional matrix systems, then with mixtures
 thereof. A mixed system was explored because of its proven potential to ionize certain analytes
 more effectively. I After several mixtures were tested, an optimized system was chosen.
- The optimized mixture consisted of approximately equal parts of 2,5 dihydroxybenzoic
 acid (2,5 DHB), dithranol, and a-cyano-4-hydroxy-cinnamic acid (HCCA). Previous work with
 phospholipids indicated that sample treatment with Nal promotes the formation of sodiated
 ions and increases the sensitivity and S/N ratio of the measurement. 2 Sodium lodide was added
 to the mixed and individual matrix systems to determine its effect on ionization.
- A mixture of the aforementioned compounds was then analyzed to ensure ionization of all of the components. The data from the optimized system mixture analysis was also compared with traditional matrices.
- To confirm peak identities, TOF-TOF collision-induced dissociation experiments were performed.




Figure 2. Shown above is the spectrum of the analysis of a seven-component mixture with the optimized Nal-doped matrix. The components (listed above) were all successfully sodiated and analyzed.

Conclusions

- With the use of an optimized mixed matrix system, it is possible to characterize trace biodiesel components in one rapid analysis.
- The use of MALDI-TOF MS circumvents complications arising from complex separation techniques and has a high throughput, allowing many samples to be analyzed and compared in one experiment.
- The mixed optimized matrix when doped with Nal has a higher sensitivity than doped and undoped pure matrices.
- · Tandem MS can be employed in complex mixtures to confirm peak identities

¢----

- S. Laugesen and P. Roepstorff, Combination of Two Matrices Results in Improved Performance of MALDI-MS for Peptide Mass Mapping and Protein Analysis, Journal of the American Society for Mass Spectrometry, 2003. 14, 992-1002.
- (2) Y. Isahida, A.J. Madonna, J.C. Rees, M.A. Meetani and K.J. Voorhees, Discrimination of Enterobacteriaceae based on Spectral Patterns of Phospholipids using MALDI-TOF/MS Combined with On-Probe Sample Pretreatment, Rapid Commun. in Mass Spectrom., 2002, 16, 1877.

Acknowledgement

The authors thank the U.S. Department of Energy Vehicle Technologies Program for project funding and support. We also thank the University of Colorado School of Pharmacy Mass Spectrometry Facility for analyzing samples.