

**NISTIR 1149-01**

# **NIST Time and Frequency Bulletin**

Petrina C. Potts, Editor

<http://dx.doi.org/10.6028/NIST.IR.1149-01>

**NISTIR 1149-01**

# **NIST Time and Frequency Bulletin**

Petrina C. Potts, Editor  
*Time and Frequency Division*  
*Physical Measurement Laboratory*

<http://dx.doi.org/10.6028/NIST.IR.1149-01>

January 2013



U.S. Department of Commerce  
*Rebecca Blank, Acting Secretary*

National Institute of Standards and Technology  
*Patrick D. Gallagher, Under Secretary of Commerce for Standards and Technology and Director*

**NIST TIME AND FREQUENCY BULLETIN**  
**NIST IR 1149-01**

**No. 661 January 2013**

1. GENERAL BACKGROUND INFORMATION.....	2
2. TIME SCALE INFORMATION .....	2
3. BROADCAST OUTAGES OVER FIVE MINUTES AND WWVB PHASE PERTURBATIONS.....	4
4. NOTES ON NIST TIME SCALES AND PRIMARY STANDARDS.....	4
5. UTC(NIST) – AT1 PARAMETERS .....	5

This bulletin is published monthly. Address correspondence to:

Petrina C. Potts, Editor  
Time and Frequency Division  
National Institute of Standards and Technology  
325 Broadway  
Boulder, CO 80305-3328  
(303) 497-3295  
Email: [ppotts@boulder.nist.gov](mailto:ppotts@boulder.nist.gov)



---

U.S. DEPARTMENT OF COMMERCE, Rebecca Blank, Acting Secretary  
NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY, Patrick D. Gallagher, Under Secretary of Commerce  
for Standards and Technology and Director

## 1. GENERAL BACKGROUND INFORMATION

### ACRONYMS AND ABBREVIATIONS USED IN THIS BULLETIN

ACTS	- Automated Computer Time Service		
BIPM	- Bureau International des Poids et Mesures		
CS	- Cesium Standard		
GPS	- Global Positioning System		
IERS	- International Earth Rotation Service		
LORAN	- Long Range Navigation		
MC	- Master Clock		
MJD	- Modified Julian Date		
NIST	- National Institute of Standards and Technology		
NOAA	- National Oceanic and Atmospheric Administration		
NVLAP	- National Voluntary Laboratory Accreditation Program	ns	- nanosecond
SI	- International System of Units	μs	- microsecond
TA	- Atomic Time	ms	- millisecond
TAI	- International Atomic Time	s	- second
USNO	- United States Naval Observatory	min	- minute
UTC	- Coordinated Universal Time		

## 2. TIME SCALE INFORMATION

The values listed below are based on data from the IERS, the USNO, and NIST. The UTC(USNO,MC) - UTC(NIST) values are averaged measurements from all available common-view GPS satellites (see bibliography on page 5). UTC - UTC(NIST) data are on page 3.

0000 HOURS COORDINATED UNIVERSAL TIME			
DEC 2012	MJD	UT1-UTC(NIST) (±5 ms)	UTC(USNO,MC) - UTC(NIST) (±20 ns)
6	56267	+300 ms	-2 ns
13	56274	+291 ms	-2 ns*
20	56281	+284 ms	-3 ns
27	56288	-280 ms	-4 ns

\*This value was interpolated due to a receiver system failure

The master clock pulses used by the WWV, WWVH, and WWVB time-code transmissions are referenced to the UTC(NIST) time scale. Occasionally, 1 s is added to the UTC time scale. This second is called a leap second. Its purpose is to keep the UTC time scale within ±0.9 s of the UT1 astronomical time scale, which changes slightly due to variations in the Earth's period of rotation.

**NOTE:** No leap second will be added at the end of December 2012.

Positive leap seconds, beginning at 23 h 59 min 60 s UTC and ending at 0 h 0 min 0 s UTC, were inserted in the UTC time scale on 30 June 1972, 1981-1983, 1985, 1992-1994, 1997, and 2012, and on 31 December 1972-1979, 1987, 1989, 1990, 1995, 1998, 2005, and 2008.

The use of leap seconds ensures that UT1 - UTC will always be held within ±0.9 s. The current value of UT1 - UTC is called the DUT1 correction. DUT1 corrections are broadcast by WWV, WWVH, WWVB, and ACTS and are printed below. These corrections may be added to received UTC time signals in order to obtain UT1.

DUT1 = UT1 - UTC =	+0.2 s beginning 0000 UTC 31 January 2013
	+0.3 s beginning 0000 UTC 25 October 2012
	+0.4 s beginning 0000 UTC 01 July 2012
	- 0.6 s beginning 0000 UTC 10 May 2012
	- 0.5 s beginning 0000 UTC 09 February 2012
	- 0.4 s beginning 0000 UTC 04 November 2011
	- 0.3 s beginning 0000 UTC 12 May 2011
	- 0.2 s beginning 0000 UTC 06 January 2011
	- 0.1 s beginning 0000 UTC 03 June 2010
	+0.0 s beginning 0000 UTC 11 March 2010
	+0.1 s beginning 0000 UTC 12 November 2009
	+0.2 s beginning 0000 UTC 11 June 2009
	+0.3 s beginning 0000 UTC 12 March 2009

The difference between UTC(NIST) and UTC has been within  $\pm 100$  ns since July 6, 1994. The table below shows values of UTC - UTC(NIST) as supplied by the BIPM in their *Circular T* publication for the most recent 310-day period in which data are available. Data are given at ten-day intervals. Five-day interval data are available in *Circular T*.

0000 Hours Coordinated Universal Time		
DATE	MJD	UTC-UTC(NIST) ns
Nov. 28, 2012	56259	-3.1
Nov. 18, 2012	56249	-1.3
Nov. 8, 2012	56239	-1.4
Oct. 29, 2012	56229	0.5
Oct. 19, 2012	56219	2.5
Oct. 9, 2012	56209	2.8
Sep. 29, 2012	56199	2.9
Sep. 19, 2012	56189	3.9
Sep. 9, 2012	56179	3.3
Aug. 30, 2012	56169	2.8
Aug. 20, 2012	56159	2.7
Aug. 10, 2012	56149	2.8
Jul. 31, 2012	56139	2.3
Jul. 21, 2012	56129	2.9
Jul. 11, 2012	56119	3.5
Jul. 1, 2012	56109	4.8
Jun. 21, 2012	56099	5.5
Jun. 11, 2012	56089	7.8
Jun. 1, 2012	56079	9.3
May 22, 2012	56069	10.2
May 12, 2012	56059	9.1
May 2, 2012	56049	7.0
Apr. 22, 2012	56039	4.1
Apr. 12, 2012	56029	0.9
Apr. 2, 2012	56019	-0.6
Mar. 23, 2012	56009	-1.9
Mar. 13, 2012	55999	-2.3
Mar. 3, 2012	55989	-3.4
Feb. 22, 2012	55979	-4.7
Feb. 12, 2012	55969	-2.7
Feb. 2, 2012	55959	-0.8

### 3. BROADCAST OUTAGES OVER FIVE MINUTES AND WWVB PHASE PERTURBATIONS

OUTAGES OF 5 MINUTES OR MORE						PHASE PERTURBATIONS 2 ms			
Station	Dec 2012	MJD	Began UTC	Ended UTC	Freq.	Dec 2012	MJD	Began UTC	End UTC
WWVB	12-09-12	56270	0323	0400	60 kHz				
WWV									
WWVH									

### 4. NOTES ON NIST TIME SCALES AND PRIMARY STANDARDS

Primary frequency standards developed and operated by NIST are used to provide accuracy (rate) input to the BIPM. NIST-F1, a cold-atom cesium fountain frequency standard, has served as the U.S. primary standard of time and frequency since 1999. The uncertainty of NIST-F1 is currently about 3 parts in  $10^{16}$ .

The AT1 scale is run in real-time by use of data from an ensemble of cesium standards and hydrogen masers. It is a free-running scale whose frequency is maintained as nearly constant as possible by choosing the optimum weight for each clock that contributes to the computation.

UTC(NIST) is generated as an offset from our real-time scale AT1. It is steered in frequency towards UTC by use of data published by the BIPM in its *Circular T*. Changes in the steering frequency will be made, if necessary, at 0000 UTC on the first day of the month, and occasionally at mid-month. A change in frequency is limited to no more than  $\pm 2$  ns/day. The frequency of UTC(NIST) is kept as stable as possible at other times.

UTC is generated at the BIPM by use of a post-processed time-scale algorithm and is not available in real-time. The parameters that we use to generate UTC(NIST) in real-time are therefore based on an extrapolation of UTC from the most recent available data.

#### References:

Allan, D.W.; Hellwig, H.; and Glaze, D.J., "An accuracy algorithm for an atomic time scale," *Metrologia*, Vol.11, No.3, pp. 133-138 (1975).

Allan, D.W.; Davis, D.D.; Weiss, M.A.; Clements, A.; Guinot, B.; Granveaud, M.; Dorenwendt, K.; Fischer, B.; Hetzel, P.; Aoki, S.; Fujimoto, M.; Charron, L.; and Ashby, N., "Accuracy of international time and frequency comparisons via global positioning system satellites in common-view," *IEEE Transactions on Instrumentation and Measurement*, Vol. IM-34, pp.118-125 (1985).

Heavner, T.P.; Jefferts, S.R.; Donley, E.A.; Shirley, J.H. and Parker, T.E., "NIST F1; recent improvements and accuracy evaluations," *Metrologia*, Vol. 42, pp. 411-422 (2005).

Jefferts, S.R.; Shirley, J.; Parker, T.E.; Heavner, T.P.; Meekhof, D.M.; Nelson, C.; Levi, F.; Costanza, G.; De Marchi, A.; Drullinger, R.; Hollberg, L.; Lee, W.D.; and Walls, F.L., "Accuracy evaluation of NIST-F1," *Metrologia*, Vol. 39, pp. 321-336 (2002).

Lewandowski, W. and Thomas, C., "GPS Time transfer," *Proceedings of the IEEE*, Vol. 79, pp. 991-1000 (1991).

Parker, T.E.; Jefferts, S.R.; Heavner, T.P.; and Donley, E.A., "Operation of the NIST-F1 caesium fountain primary frequency standard with a maser ensemble, including the impact of frequency transfer noise," *Metrologia*, Vol. 42, pp. 423-430 (2005).

Weiss, M.A.; Allan, D.W., "An NBS Calibration Procedure for Providing Time and Frequency at a Remote Site by Weighting and Smoothing of GPS Common View Data," *IEEE Transactions on Instrumentation and Measurement*, Vol. IM-36, pp. 572-578 (1987).

## 5. UTC(NIST) – AT1 PARAMETERS

The table below lists parameters that are used to define UTC(NIST) with respect to our real-time scale AT1. To find the value of UTC(NIST) - AT1 at any time T (expressed as a Modified Julian Day, including a fraction if needed), the appropriate equation to use is the one for which the desired T is greater than or equal to the entry in the  $T_0$  column and less than the entry in the last column. The values of  $x_{ls}$ ,  $x$ , and  $y$  for that month are then used in the equation below to find the desired value. The parameters  $x$  and  $y$  represent the offsets in time and frequency, respectively, between UTC(NIST) and AT1; the parameter  $x_{ls}$  is the number of leap seconds applied to both UTC(NIST) and UTC, as specified by the IERS. Leap seconds are not applied to AT1.

UTC(NIST) - AT1 = $x_{ls} + x + y*(T - T_0)$					
Month	$x_{ls}$ (s)	$x$ (ns)	$y$ (ns/d)	$T_0$ (MJD)	Valid until 0000 on: (MJD)
Feb 13	-35	386469.5	-37.7*	56324	56352
Jan 13	-35	385300.8	-37.7	56293	56324*
Dec 12	-35	-384132.1	-37.7	56262	56293
Nov 12	-35	-383001.1	-37.7	56232	56262
Oct 12	-35	-381832.4	-37.7	56201	56232
Sep 12	-35	-380701.4	-37.7	56171	56201
Aug 12	-35	-379532.7	-37.7	56140	56171
Jul 12	<b>-35</b>	-3783640	-37.7	56109	56140
Jun 12	-34	-377233	-37.7	56079	56109
May 12	-34	-376705.2	-37.7	56065	56079
May 12	-34	-376059.2	-38	56048	56065†
Apr 12	-34	-374919.2	-38	56018	56048
Mar 12	-34	-373741.2	-38	55987	56018
Feb 12	-34	-373399.2	-38	55978	55987
Feb 12	-34	-372643.2	-37.8	55958	55978†
Jan 12	-34	-371471.4	37.8	55927	55958
Dec 11	-34	-370293.4	-38.0	55896	55927
Nov 11	-34	-370027.4	-38.0	55889	55896
Nov 11	-34	-369158	-37.8	55866	55889†
Oct 11	-34	-368477.6	-37.8	55848	55866
Oct 11	-34	-367983.6	-38.0	55835	55848†

† Rate change in mid-month

\*Provisional value