

RESULTS OF THE U.S. GEOLOGICAL SURVEY'S SECOND INTERNATIONAL
INTERLABORATORY ANALYTICAL COMPARISON STUDY--STANDARD REFERENCE
WATER SAMPLES M-86 (MAJOR CONSTITUENTS), T-87 (TRACE CONSTITUENTS), AND
P-5 (PRECIPITATION SNOWMELT)

By Victor J. Janzer

U.S. GEOLOGICAL SURVEY

Water-Resources Investigations Report 85-4049



Denver, Colorado

1985

UNITED STATES DEPARTMENT OF THE INTERIOR

DONALD PAUL HODEL, Secretary

GEOLOGICAL SURVEY

Dallas Peck, Director

For additional information write to:

Victor J. Janzer
U.S. Geological Survey
Water Resources Division
Mail Stop 407, Box 25046
Denver Federal Center
Denver, Colorado 80225
Telephone: (303)236-1924
FTS: 776-1924

Copies of this report can be purchased from:

Open-File Services Section
Western Distribution Branch
U.S. Geological Survey
Box 25425, Federal Center
Denver, Colorado 80225
Telephone: (303)236-7476
FTS: 776-7476

ERRATA SHEET FOR WRIR 85-4049

Table 11

Footnote 1/ should read: Except specific conductance (microsiemens or micromhos per centimeter at 25°C); pH (units); boron, bromide, iodide, strontium, and vanadium (micrograms per liter).

Table 12

Mean Concentration heading should show footnote 1/ Except acidity (milligrams per liter).

CONTENTS

	<u>Page</u>
Abstract	1
Introduction	2
Purpose and plan	2
Preparation of samples	2
Determinations	4
Standard Reference Water Sample M-86 (major constituents)	4
Standard Reference Water Sample T-87 (trace constituents)	4
Standard Reference Water Sample P-5 (precipitation snowmelt)	4
Laboratory performance and reported values	5
Statistical evaluation	6
Participating laboratories	7
References	9

TABLES

Table 1. Explanation of abbreviations, symbols, and terms used in computer-printout parts of subsequent tables..	10
2. Overall laboratory performance; SRWS M-86 (major constituents)	11
3. Overall laboratory performance; SRWS T-87 (trace constituents)	15
4. Overall laboratory performance; SRWS P-5 (precipitation snowmelt)	18
5. Analytical data; SRWS M-86 (major constituents)	20
6. Statistics by method for SRWS M-86 (major constituents)	40
7. Analytical data; SRWS T-87 (trace constituents)	44
8. Statistics by method for SRWS T-87 (trace constituents)	66
9. Analytical data; SRWS P-5 (precipitation snowmelt)	70
10. Statistics by method for SRWS P-5 (precipitation snowmelt)	81
11. Comparison of domestic and international laboratory analyses for SRWS M-86 (major constituents)	83
12. Comparison of domestic and international laboratory analyses for SRWS T-87 (trace constituents)	84
13. Comparison of domestic and international laboratory analyses for SRWS P-5 (precipitation snowmelt)	85

RESULTS OF THE U.S. GEOLOGICAL SURVEY'S SECOND INTERNATIONAL
INTERLABORATORY ANALYTICAL COMPARISON STUDY--STANDARD REFERENCE
WATER SAMPLES M-86 (MAJOR CONSTITUENTS), T-87 (TRACE
CONSTITUENTS), AND P-5 (PRECIPITATION SNOWMELT)

By Victor J. Janzer

ABSTRACT

The U.S. Geological Survey began an interlaboratory testing program of Standard Reference Water Samples in 1962. Program objectives have been to provide a means for participating analytical laboratories to: (1) Identify analytical problems, (2) ascertain the accuracy and precision of common water analyses and analytical methods, and (3) obtain reference samples for continuing quality-assurance testing. Participation in this continuing quality-assurance program is mandatory for all domestic laboratories providing water-analysis data for Survey use and storage in the WATSTORE data storage system, if appropriate Standard Reference Water Samples are available.

The program was expanded in October 1982 to include international laboratories. This report presents analytical data submitted by the 53 laboratories in other countries that analyzed the reference samples distributed in October 1983. Statistical evaluation of the data and performance ratings achieved by the laboratories for each determination are given in nine tables. Comparisons of the most probable values for the constituents determined by both international and domestic laboratories are also presented.

INTRODUCTION

The U.S. Geological Survey began an interlaboratory testing program of standard reference water samples in 1962. Principal purposes of the program have been to provide a means for participating domestic analytical laboratories to: (1) Identify analytical problems, (2) ascertain the accuracy and precision of the analytical methods for determining the various constituents and physical properties of water and (3) obtain reference samples for continuing quality-assurance testing. Twenty-three Geological Survey laboratories participated in the 1962 effort to determine six constituents in a single standard reference water sample (SRWS) containing major constituents. Now, more than 120 domestic laboratories participate in the program that uses as many as 8 SRWS types; major constituents, trace constituents, nutrients, herbicides, insecticides, water-sediment mixture for trace metals, precipitation snowmelt, and priority pollutants.

Participation in this continuing quality-assurance program is mandatory for all domestic laboratories providing water-analysis data for Survey use and storage in the WATSTORE data-storage system, if appropriate SRWS are available. Major constituent, trace constituent, and nutrient SRWS are prepared and distributed to domestic laboratories twice each year. One or more of the other SRWS types may also be included.

The program was expanded in October 1982 to include a number of international laboratories. Samples have been distributed to the laboratories in other countries only once a year, during October of 1982 and 1983. This report summarizes analytical data submitted by the 53 international laboratories that analyzed the reference samples distributed in October 1983. Statistical evaluation of the data and performance ratings achieved by the laboratories for each determination are given in nine tables. Comparison of the most probable values for constituents determined by both international and domestic laboratories are also presented in three additional tables. The domestic data were reported previously (Janzer and Latal, 1984).

PURPOSE AND PLAN

As a means of providing an independent, objective evaluation of the water-quality data published by the U.S. Geological Survey, SRWS are prepared and distributed for analysis at regular intervals. SRWS M-86 (major constituents), T-87 (trace constituents), and P-5 (precipitation snowmelt) were distributed to 100 domestic laboratories in October 1983. In addition, SRWS were sent to 68 international laboratories that indicated their willingness to analyze these reference samples. All samples are not analyzed by all laboratories nor do all laboratories participate in each round of analyses.

Each laboratory was requested to indicate the analytical methods used and to perform at least those determinations that it makes routinely. Laboratories participating in this study are identified only by confidential code numbers.

PREPARATION OF SAMPLES

SRWS M-86 (major constituents), and T-87 (trace constituents) were each prepared from a surface water collected from the same source. Samples were filtered through a 5- μm (micrometer) nominal size prefilter and a 0.45- μm membrane filter into a 1325-L (liter) polyethylene drum. Thymol, about 1.25 mg/L (milligrams per liter), was added to SRWS M-86 and T-87, to reduce growth of fungus.

Some trace constituents (vanadium and fluoride) were added to SRWS M-86. No constituent additions were made to SRWS T-87 but it was acidified to a pH of about 1.5 with nitric acid. Each sample was mixed overnight with a motor-driven, Teflon^{1/}-coated stirrer, filtered through a 0.45-μm membrane filter, and passed through a flow-through 254-nm ultraviolet sterilizer and bottled, under ultraviolet radiation, in 1-L autoclaved polypropylene bottles or dry-heat sterilized Teflon bottles.

SRWS P-5 (precipitation snowmelt) was prepared by melting snow collected in several 200-L polyethylene drums. After melting, the sample was filtered through a 0.45-μm membrane filter. No additions of any kind were made to this sample. After mixing overnight, the sample was again filtered through a 0.45-μm filter, sterilized by passage through the flow-through ultraviolet sterilizer and bottled in 1-L autoclaved polypropylene bottles or dry-heat sterilized Teflon bottles under ultraviolet radiation.

^{1/}The use of the trade name in this report is for identification purposes only and does not constitute endorsement by the U.S. Geological Survey.

DETERMINATIONS

Determinations for each of the SRWS and their abbreviations are listed below.

SRWS M-86 (major constituents) (results in milligrams per liter^{1/})

ALK(CACO ₃)	= Alkalinity (as CaCO ₃)	NA	= Sodium
B	= Boron	NO ₂ -N	= Nitrite as nitrogen
BR	= Bromide	NO ₃ -N	= Nitrate as nitrogen
CA	= Calcium	P, TOTAL	= Phosphorus, total as phosphorus
CL	= Chloride	PH	= pH
DSRD 180	= Dissolved solids	SIO ₂	= Silica
F	= Fluoride	SO ₄	= Sulfate
I	= Iodide	SP. COND.	= Specific conductance
K	= Potassium	SR	= Strontium
MG	= Magnesium	V	= Vanadium

SRWS T-87 (trace constituents) (results in micrograms per liter^{2/})

ACID@CACO ₃	= Acidity (as CaCO ₃)	HG	= Mercury
AG	= Silver	LI	= Lithium
AL	= Aluminum	MN	= Manganese
AS	= Arsenic	MO	= Molybdenum
BA	= Barium	NI	= Nickel
BE	= Beryllium	PB	= Lead
CD	= Cadmium	SB	= Antimony
CO	= Cobalt	SE	= Selenium
CR, TOTAL	= Chromium, total	SR	= Strontium
CU	= Copper	TL	= Thallium
FE	= Iron	ZN	= Zinc

SRWS P-5 (precipitation snowmelt) (results in milligrams per liter^{3/})

CA	= Calcium	NH ₃ -N	= Ammonia as nitrogen
CL	= Chloride	NO ₃ -N	= Nitrate as nitrogen
F	= Fluoride	PH	= pH
K	= Potassium	SO ₄	= Sulfate
MG	= Magnesium	SP. COND.	= Specific conductance
NA	= Sodium		

^{1/} Except specific conductance (microsiemens or micromhos per centimeter at 25° C); pH (units); boron, bromide, iodide, strontium, and vanadium (micrograms per liter).

^{2/} Except acidity (milligrams per liter).

^{3/} Except pH (units) and specific conductance (microsiemens or micromhos per centimeter at 25° C).

LABORATORY PERFORMANCE AND REPORTED VALUES

To facilitate interlaboratory performance comparisons, ratings based on the analyses of each SRWS are included in this report as tables 2-4 (all tables are at back of report; the abbreviations and symbols used in the tables are defined in table 1). Laboratory performance for each constituent is rated on an arbitrary scale of 0 to 4 based on the number of "standard deviations" from the mean determined for each constituent as indicated below:

4 (Excellent)	0.00 to 0.50 standard deviations
3 (Good)	0.51 to 1.00 standard deviations
2 (Satisfactory)	1.01 to 1.50 standard deviations
1 (Questionable)	1.51 to 2.00 standard deviations
0 (Poor)	Greater than 2.00 standard deviations

When the analyses for a constituent are extremely precise, these ratings may be overly severe and should be considered only as indicators of relative performance. Averages of the constituent ratings for each SRWS are given for each laboratory in tables 2-4 of overall laboratory performance.

The values reported for all constituents determined in each SRWS are listed in tables 5, 7, and 9. Each value has been rounded off, when necessary, to conform to U.S. Geological Survey policy on reporting analytical data. Laboratories were requested to indicate the general method used for each determination. When this information was provided, method identifications have been included with the analytical data. Statistical information by method for each determination are listed in tables 6, 8, and 10. Summary comparisons of other-country and domestic laboratory analyses of SRWS M-86, T-87, and P-5 are presented in tables 11-13. Mean concentrations for most constituents determined by both laboratory groups show good agreement.

Participants are encouraged to submit comments or suggestions concerning this program to:

Victor J. Janzer
U.S. Geological Survey
5293 Ward Road
Arvada, Colorado 80002
U.S.A.

STATISTICAL EVALUATION

A statistical evaluation of the data was made to estimate the most probable value (MPV) for each of the constituents. Values reported as "less than" were considered "not determined", and were not used in the computation of the means, standard deviations, and so forth. These data are indicated as "ignored" in the computer listings.

The mean, standard deviation, and confidence limits about the mean are usually reported to one more significant figure than the reported value. Statistical information is tabulated for each method used by three or more laboratories to determine a specific constituent. Tables giving the mean and standard deviation determined by that method, and the number of laboratories which used it, follow the analytical data tables for each SRWS.

Outliers in each data set are identified and rejected based on the T values (Grubbs' test) described and tabulated in the American Society for Testing and Materials (1981) Recommended Practice E-178 (1980). If the computed T value is greater than the tabulated value for the number of samples and the significance level selected, the outlier is rejected. T is computed by:

$$T = \frac{X_n - \bar{X}}{S}$$

where T = T value for probable outlier,

X_n = concentration of probable outlier,

\bar{X} = arithmetic mean (average) of all values, and

S = standard deviation of all values.

After rejection of the outliers, the data remaining for each constituent were used to calculate the means, standard deviations and percent deviation from the mean. Values identified as outliers were omitted when calculating the means and standard deviations for each determination listed by "method". The total range for each constituent included the outliers. Confidence intervals about the mean were also calculated. These define the range within which the true value is expected to occur with a confidence level of 95 percent.

PARTICIPATING LABORATORIES

AUSTRALIA, Brisbane: Government Chemical Laboratory

AUSTRALIA (South), Eastwood: The Australian Mineral Development Laboratory

AUSTRALIA (Western), Perth: Government Chemical Laboratories

AUSTRIA, Vienna: Isotope Hydrology, International Atomic Energy Agency

BRAZIL, Minas Gerais: Fundacao Centro Technologico de Minas Gerais

BRAZIL, Minas Gerais: Companhia de Pesquisa de Recursos Minerais

CANADA, Calgary, Alberta: Inland Water Directorate, Western Region Water Quality Branch

CANADA, West Vancouver, BC: EPS-DOE Laboratory Services

CANADA, Winnipeg, Manitoba: Technical Services Laboratory

CANADA, Ottawa, Ontario: Energy, Mines & Resources Canada, Geological Survey of Canada

CANADA, Rexdale, Ontario: Acid Precipitation Studies, Ontario Ministry of Environment

CANADA, Rexdale, Ontario: Rivers & Lakes Laboratory, Ontario Ministry of Environment

CANADA, Toronto, Ontario: Analytical Services Section, Ontario Hydrology

COLOMBIA, Bogota: Instituto Colombiano de Hidrologia, Ministry of Agriculture

CZECHOSLOVAKIA, Bratislava: Institute of Geology, Department of Hydrogeochemistry

CZECHOSLOVAKIA, Praha: Geological Survey Prague, Chemical Laboratory

CZECHOSLOVAKIA, Zilina: IGHP, Hydrochemical Laboratory

ENGLAND, London: Water & Wastewater Subdivision, Laboratory of the Government Chemist

ENGLAND, Wallingford, Oxfordshire: Institute of Geological Sciences, Hydrogeological Department

FEDERAL REPUBLIC OF WEST GERMANY, Koblenz: Bundesanstalt fur Gewasserkunde

FEDERAL REPUBLIC OF WEST GERMANY, Neuhof: Chemische und Biologische Laboratorien GmbH

FINLAND, Helsinki: National Board of Waters, Research Laboratory

GREECE, Athens: Soil & Water Laboratory, Hellenic Republic Ministry of Agriculture

HUNGARY, Budapest: Hungarian Geological Survey

HUNGARY, Budapest: VITUKI, Research Centre for Water Resources Development

INDIA, Lucknow (UP): Central Ground Water Board, Northern Region

ISRAEL, Jerusalem: Emission Spectrometric Laboratory, Geological Survey of Israel

ISRAEL, Jerusalem: Hydrological Service Water Commission, Ministry of Agriculture

ISRAEL, Tel-Aviv: Tahal Consulting Engineers, Ltd.

ITALY, Venezia: Universita Degli Studi Di Venezia, Istituto di Chimica Generale Ed Inorganica

JORDAN, Amman: Natural Resources Authority, Water & Isotope Laboratory

NEW ZEALAND, Lower Hutt, Petone: Department Scientific and Industrial Research

NORWAY, Oslo: Norwegian Institute for Water Research

PORTUGAL, Lisbon: Universidade Nova De Lisboa, Department Environmental Engineering

REPUBLIC OF CHINA, Taipei, Taiwan: Water Resource Planning Commission, Ministry of Economic Affairs

SAUDI ARABIA, Abqaiq: ARAMCO, Abqaiq Laboratory

SOUTH AFRICA, Bellville: National Institute for Water Research, CSIR, Cape Regional Laboratory

SOUTH AFRICA, Cape Town: City of Cape Town, City Engineer's Department

SOUTH AFRICA, Germiston: Johannesburg Consolidated Investment Co., Ltd, Minerals Processing Research Laboratory

SOUTH AFRICA, Johannesburg: McLachlan & Lazar (PTY) Ltd.

SOUTH AFRICA, Natal: National Institute for Water Research, CSIR, Natal Regional Office

SOUTH AFRICA, Pretoria: National Institute for Water Research, CSIR

SOUTH AFRICA, Pretoria: Hydrological Research Institute, Department of Environment Affairs

SULTANATE OF OMAN, Ruwi: Public Authority Water Resources

SWEDEN, Norrkoping: Sveriges meteorologiska och hydrologiska institut

SWEDEN, Solna: National Swedish Environment Protection Board, Research Laboratory

SWEDEN, Uppsala: Water Quality Laboratory, Statens Naturvardsverk, The National Swedish Environmental Protection Board

SWITZERLAND, Dubendorf: EAWAG

TANZANIA, Dar es Salaam: Project Preparation Division

THE NETHERLANDS, Lelystad: Governmental Institute for Sewage & Wastewater Treatment (RIZA)

THE NETHERLANDS, Oosterzee: Limnologische Institute, Tjeukemeer Laboratory

USSR, Leningrad: VSEGEI

ZIMBABWE, Harare: City of Harare, Department of Works

REFERENCES

- American Society for Testing and Materials, 1981, Annual Book of ASTM Standards, Part 41: Philadelphia, 1390 p.
- _____, 1982, Annual book of ASTM standards, Part 31: Philadelphia, 1554 p.
- Janzer, V. J., and Latal, K. A., 1984, Report of the U.S. Geological Survey's analytical evaluation program--standard reference water samples M-86 (major constituents), T-87 (trace constituents), N-10 and N-11 (nutrients), P-5 (precipitation snowmelt) and POL-1 (priority pollutants): U.S. Geological Survey Open-File Report 84-128, 140 p.

Table 1---Explanation of abbreviations and symbols used in computer-printout parts of subsequent tables

APDC/MIBK - ammonium pyrrolidine dithiocarbamate/methyl isobutyl ketone
AUTO - automated
BLK - block
DEV - deviation
DIG - digestion
EDTA - ethylenediaminetetraacetic acid
H₂SO₄ - sulfuric acid
IGNORED - values reported as less than detection level and not used in statistical analyses
INTRVL - interval
K & HG SO₄ - potassium & mercuric sulfate
PCT - percent
PDCA/CHCl₃ - pyrrolidine dithiocarbamic acid/chloroform
REJECT - values identified as an outlier and not used in statistical analyses
SRWS - standard reference water sample
STD - standard

TABLE 2.—OVERALL LABORATORY PERFORMANCE

8000S M-86 (MAJUN-INTEGRATIONAL)

NR = NOT RATED

ND = NOT DETERMINED

LT = LESS-THAN VALUE REPORTED, NOT RATED

N = NUMBER OF CONSTITUENTS LABORATORY DETERMINED

AVG = AVERAGE LABORATORY PERFORMANCE RATING

RATING	4 (EXCELLENT)	5 (GOOD)	2 (SATISFACTORY)	1 (QUESTIONABLE)	0 (POOR)
LAB ALA (CACUS) B	ND	3	2	1	ND
002	4	4	3	1	ND
003	4	4	3	1	ND
004	4	4	3	1	ND
005	5	5	3	1	ND
007	5	5	3	1	ND
008	5	5	3	1	ND
009	4	4	3	1	ND
010	4	4	3	1	ND
011	4	4	3	1	ND
012	4	4	3	1	ND
013	ND	1	0	0	ND
014	ND	1	0	0	ND
015	0	0	0	0	ND
017	1	1	0	0	ND
018	ND	ND	ND	ND	ND
019	0	0	0	0	ND
020	0	0	0	0	ND
022	0	0	0	0	ND
023	0	0	0	0	ND
025	1	1	0	0	ND
026	3	4	0	0	ND
027	5	5	0	0	ND
030	ND	ND	ND	ND	ND
032	4	4	0	0	ND
033	2	2	0	0	ND
034	ND	2	0	0	ND
035	2	2	0	0	ND
036	ND	ND	0	0	ND
037	0	0	2	0	ND
038	0	0	0	2	ND
039	4	ND	ND	ND	ND
040	ND	4	ND	ND	ND
041	4	4	1	0	ND
042	4	4	1	0	ND
043	4	4	1	0	ND
045	4	4	1	0	ND
047	4	4	1	0	ND
048	3	4	1	0	ND
053	3	4	1	0	ND
055	4	ND	ND	ND	ND
057	ND	ND	ND	ND	ND

TABLE 2.—OVERALL LABORATORY PERFORMANCE
SKWS W-R6 (MAJUK-INTERNATIONAL)

RATING	LAB	NO.3-N		P. TOTAL		PH		SI02		SUM		SP.		COND.		SR	
		ND	NU	ND	NU	ND	NU	ND	NU	ND	NU	ND	NU	ND	NU	ND	NU
4 (EXCELLENT)	0.90	10	0.50	SIU.	DEV.	ND	0	0	0	ND	0	N	ND	0	0	0	Avg.
3 (GOOD)	0.51	10	1.00	SIU.	DEV.	ND	0	0	0	ND	0	11	19	3.09	3.09	3.09	3.09
2 (SATISFACTORY)	1.01	10	1.50	SIU.	DEV.	ND	0	0	0	ND	0	11	19	2.47	2.47	2.47	2.47
1 (QUESTIONABLE)	1.51	10	2.00	SIU.	DEV.	ND	0	0	0	ND	0	11	19	1.65	1.65	1.65	1.65
0 (POOR)			> 2.00	SIU.	DEV.	ND	0	0	0	ND	0	11	19	2.95	2.95	2.95	2.95
						ND	0	0	0	ND	0	11	19	1.45	1.45	1.45	1.45
						ND	0	0	0	ND	0	13	19	2.77	2.77	2.77	2.77
						ND	0	0	0	ND	0	12	19	1.58	1.58	1.58	1.58
						ND	0	0	0	ND	0	15	19	2.60	2.60	2.60	2.60
						ND	0	0	0	ND	0	15	19	2.27	2.27	2.27	2.27
						ND	0	0	0	ND	0	17	19	2.48	2.48	2.48	2.48
						ND	0	0	0	ND	0	14	19	2.06	2.06	2.06	2.06
						ND	0	0	0	ND	0	15	19	3.07	3.07	3.07	3.07
						ND	0	0	0	ND	0	17	19	3.00	3.00	3.00	3.00
						ND	0	0	0	ND	0	15	19	2.40	2.40	2.40	2.40
						ND	0	0	0	ND	0	11	19	2.45	2.45	2.45	2.45
						ND	0	0	0	ND	0	13	19	2.44	2.44	2.44	2.44
						ND	0	0	0	ND	0	13	19	2.05	2.05	2.05	2.05
						ND	0	0	0	ND	0	13	19	3.43	3.43	3.43	3.43
						ND	0	0	0	ND	0	12	19	1.55	1.55	1.55	1.55
						ND	0	0	0	ND	0	14	19	2.97	2.97	2.97	2.97
						ND	0	0	0	ND	0	15	19	3.22	3.22	3.22	3.22
						ND	0	0	0	ND	0	9	19	3.78	3.78	3.78	3.78
						ND	0	0	0	ND	0	13	19	2.05	2.05	2.05	2.05
						ND	0	0	0	ND	0	14	19	3.07	3.07	3.07	3.07
						ND	0	0	0	ND	0	15	19	2.75	2.75	2.75	2.75
						ND	0	0	0	ND	0	16	19	2.13	2.13	2.13	2.13
						ND	0	0	0	ND	0	13	19	2.79	2.79	2.79	2.79
						ND	0	0	0	ND	0	6	19	2.77	2.77	2.77	2.77
						ND	0	0	0	ND	0	12	19	3.13	3.13	3.13	3.13
						ND	0	0	0	ND	0	4	19	2.25	2.25	2.25	2.25
						ND	0	0	0	ND	0	14	19	3.56	3.56	3.56	3.56
						ND	0	0	0	ND	0	16	19	2.91	2.91	2.91	2.91
						ND	0	0	0	ND	0	10	19	1.60	1.60	1.60	1.60
						ND	0	0	0	ND	0	13	19	3.06	3.06	3.06	3.06
						ND	0	0	0	ND	0	19	19	3.52	3.52	3.52	3.52
						ND	0	0	0	ND	0	9	19	1.67	1.67	1.67	1.67
						ND	0	0	0	ND	0	13	19	2.92	2.92	2.92	2.92

TABLE 2.—OVERALL LABORATORY PERFORMANCE

RATING	SKW6 H-66 (HAJUN-INTERNATIONAL)		HJK H-66 (HAJUN-INTERNATIONAL)		HJK H-66 (HAJUN-INTERNATIONAL)	
	EXCELLENT	GOOD	GOOD	GOOD	GOOD	GOOD
4 (EXCELLENT)	0.00	10 0.50	SIU. DEV.	SIU. DEV.	SIU. DEV.	SIU. DEV.
3 (GOOD)	0.51	10 1.00	SIU. DEV.	SIU. DEV.	SIU. DEV.	SIU. DEV.
2 (SATISFACTORY)	1.01	10 1.50	SIU. DEV.	SIU. DEV.	SIU. DEV.	SIU. DEV.
1 (QUESTIONABLE)	1.51	10 2.00	SIU. DEV.	SIU. DEV.	SIU. DEV.	SIU. DEV.
0 (POOR)	> 2.00	SIU. DEV.	SIU. DEV.	SIU. DEV.	SIU. DEV.	SIU. DEV.
AVG = AVERAGE LABORATORY PERFORMANCE RATING						
LAB	ALK (CACUS) B	BR	CA	CL	DSNU 180 F	I
058	4	ND	4	4	ND	KD
059	0	ND	4	1	0	LT
061	4	ND	ND	2	3	ND
062	2	ND	ND	0	0	ND
065	4	ND	ND	2	4	0
066	3	ND	ND	1	2	2
	1	ND	ND	3	ND	1
					ND	3
					ND	4

NN = NOT DETERMINED
ND = NOT DETERMINED
LT = LESS-THAN VALUE REPORTED, NOT DETERMINED

N

= NUMBER OF CONSTITUENTS LABORATORY DETERMINED

AVG = AVERAGE LABORATORY PERFORMANCE RATING

TABLE 2.—OVERALL LABORATORY PERFORMANCE

RATING	SKINS M-86 (MAJUN-INTERNAL)			NH = NOT RATED		
	4 (EXCELLENT)	0.00	TO 0.50	SIU.	DEV.	DEV.
5 (GOOD)	0.51	TO 1.00	SIU.	DEV.	DEV.	DEV.
2 (SATISFACTORY)	1.01	TO 1.50	SIU.	DEV.	DEV.	DEV.
1 (OUTSTANDING)	1.51	TO 2.00	SIU.	DEV.	DEV.	DEV.
0 (POOR)	> 2.00	SIU.	DEV.	DEV.	DEV.	DEV.
AVG = AVERAGE LABORATORY PERFORMANCE RATING						
LAB	NA	NU2-N	NU3-N	P.	TOTAL	PH
058	4	NU	3	4	1	4
059	3	3	4	2	1	3
061	5	4	4	ND	5	4
062	NU	4	0	ND	2	4
063	1	ND	ND	0	0	0
064	4	5	4	ND	3	4
066	3	ND	6	3	4	3
						12
						3.25

NH = NOT RATED
 ND = NOT DETERMINED
 LT = LESS-THAN VALUE REPORTED, NOT RATED
 N = NUMBER OF CONSTITUENTS LABORATORY DETERMINED
 AVG = AVERAGE LABORATORY PERFORMANCE RATING

TABLE 3.--OVERALL LABORATORY PERFORMANCE
SRWS T-67 (TRACE CONSTITUENTS)

RATING	4 (EXCELLENT)	0.00	TU	0.50	STD.	DEV.	NW = NOT RATED	NO = NOT DETERMINED	LT = LESS THAN VALUE REPORTED, NOT RATED	N = NUMBER OF CONSTITUENTS LABORATORY DETERMINED	Avg = AVERAGE LABORATORY PERFORMANCE RATING
3 (GOOD)	0.51	TU	1.00	STD.	DEV.						
2 (SATISFACTORY)	1.01	TU	1.50	STD.	DEV.						
1 (QUESTIONABLE)	1.51	TU	2.00	STD.	DEV.						
0 (POOR)			> 2.00	STD.	DEV.						
LAB	ACIUSACUSAG	AL	AS	BE	CD	CU	CR	TOT	LT	LT	LT
002	ND	ND	0	ND	LT	LT	LT	LT	LT	LT	LT
003	4	LT	0	LT	LT	LT	LT	LT	LT	LT	LT
004	2	LT	0	LT	LT	LT	LT	LT	LT	LT	LT
006	2	ND	3	ND	ND	ND	ND	0	ND	ND	ND
007	1	ND	3	ND	ND	ND	ND	0	ND	ND	ND
008	0	ND	3	ND	ND	ND	ND	0	ND	ND	ND
009	0	ND	4	ND	ND	ND	ND	0	ND	ND	ND
010	3	ND	0	ND	ND	ND	ND	0	ND	ND	ND
011	NU	4	LT	ND	ND	ND	ND	0	ND	ND	ND
012	4	ND	1	ND	ND	ND	ND	0	ND	ND	ND
013	ND	3	LT	2	3	3	ND	3	LT	3	3
014	3	ND	1	ND	ND	ND	ND	0	ND	ND	ND
015	ND	3	LT	3	3	3	ND	3	ND	ND	ND
016	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
019	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
020	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
022	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
023	4	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
025	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
026	4	ND	1	ND	ND	ND	ND	0	ND	ND	ND
027	1	ND	3	ND	ND	ND	ND	0	ND	ND	ND
030	ND	ND	4	ND	ND	ND	ND	0	ND	ND	ND
032	ND	ND	3	ND	ND	ND	ND	0	ND	ND	ND
033	2	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
034	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
035	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
036	ND	ND	3	ND	ND	ND	ND	0	ND	ND	ND
037	ND	ND	3	0	ND	ND	ND	0	ND	ND	ND
039	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
040	ND	ND	1	ND	ND	ND	ND	0	ND	ND	ND
041	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
042	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
043	2	ND	2	ND	ND	ND	ND	0	ND	ND	ND
045	ND	ND	3	ND	ND	ND	ND	0	ND	ND	ND
049	2	ND	2	ND	ND	ND	ND	0	ND	ND	ND
051	3	ND	0	ND	ND	ND	ND	0	ND	ND	ND
053	4	ND	0	ND	ND	ND	ND	0	ND	ND	ND
054	4	ND	0	ND	ND	ND	ND	0	ND	ND	ND
057	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
059	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND
065	ND	ND	ND	ND	ND	ND	ND	0	ND	ND	ND

TABLE 3.—OVERALL LABORATORY PERFORMANCE
SAMS T-87 (TRACE CONSTITUENTS)

RATING	LAB	NH = NOT RATED				LT = LTSS THAN VALUE REPORTED, NOT DETERMINED				N = NUMBER OF CONSTITUENTS LABORATORY DETERMINED				AVG = AVERAGE LABORATORY PERFORMANCE RATING			
		4 (EXCELLENT)	3 (GOOD)	2 (SATISFACTORY)	1 (QUESTIONABLE)	0 (POOR)	4 (EXCELLENT)	3 (GOOD)	2 (SATISFACTORY)	1 (QUESTIONABLE)	0 (POOR)	4 (EXCELLENT)	3 (GOOD)	2 (SATISFACTORY)	1 (QUESTIONABLE)	0 (POOR)	
4 (EXCELLENT)	002	LI	ND	LI	MN	MN	ND	LI	LI	ND	ND	ND	ND	ND	ND	ND	SR
3 (GOOD)	003	LI	LT	LT	ND	ND	2	LI	LI	ND	ND	ND	ND	ND	ND	ND	4
2 (SATISFACTORY)	004	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	ND
1 (QUESTIONABLE)	006	LT	LT	2	3	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	3
0 (POOR)	007	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	1
4 (EXCELLENT)	008	3	ND	ND	ND	ND	ND	2	ND	ND	ND	ND	ND	ND	ND	ND	4
3 (GOOD)	009	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
2 (SATISFACTORY)	010	LT	3	LT	LT	LT	LT	LT	LT	LT	LT	LT	LT	LT	LT	LT	ND
1 (QUESTIONABLE)	011	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
0 (POOR)	012	LI	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
4 (EXCELLENT)	013	3	LT	3	ND	ND	3	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
3 (GOOD)	014	3	ND	3	ND	ND	0	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
2 (SATISFACTORY)	015	4	ND	ND	ND	ND	3	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
1 (QUESTIONABLE)	016	ND	4	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
0 (POOR)	019	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
4 (EXCELLENT)	020	0	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
3 (GOOD)	022	3	ND	ND	ND	ND	ND	3	ND	ND	ND	ND	ND	ND	ND	ND	ND
2 (SATISFACTORY)	023	2	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
1 (QUESTIONABLE)	025	LT	4	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
0 (POOR)	026	0	ND	ND	ND	ND	ND	0	ND	ND	ND	ND	ND	ND	ND	ND	ND
4 (EXCELLENT)	027	3	ND	ND	ND	ND	ND	3	ND	ND	ND	ND	ND	ND	ND	ND	ND
3 (GOOD)	030	3	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
2 (SATISFACTORY)	032	5	ND	ND	ND	ND	ND	4	ND	ND	ND	ND	ND	ND	ND	ND	ND
1 (QUESTIONABLE)	033	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
0 (POOR)	034	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
4 (EXCELLENT)	035	LT	LT	LT	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
3 (GOOD)	036	4	ND	ND	ND	ND	ND	4	ND	ND	ND	ND	ND	ND	ND	ND	ND
2 (SATISFACTORY)	037	0	LT	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
1 (QUESTIONABLE)	039	LI	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
0 (POOR)	041	4	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
4 (EXCELLENT)	042	3	LI	LI	LI	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
3 (GOOD)	043	3	LI	LI	LI	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
2 (SATISFACTORY)	045	3	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
1 (QUESTIONABLE)	049	3	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
0 (POOR)	051	4	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
4 (EXCELLENT)	053	5	LT	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
3 (GOOD)	054	2	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
2 (SATISFACTORY)	057	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
1 (QUESTIONABLE)	059	LI	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
0 (POOR)	065	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND

TABLE 3.—OVERALL LABORATORY PERFORMANCE

RATING	SHWS 1-87 (TRACE CONSTITUENTS)			N	AVG.
	4 (EXCELLENT)	0.00	TO 0.50		
3 (GOOD)	0.51	TO 1.40	SID. DEV.	NH = NOT DETERMINED ND = NOT DETERMINED	
2 (SATISFACTORY)	1.01	TO 1.50	SID. DEV.	LT = LESS-THAN VALUE REPORTED, NOT RATED	
1 (QUESTIONABLE)	1.51	TO 2.00	SID. DEV.	N = NUMBER OF CONSTITUENTS LABORATORY DETERMINED	
0 (POOR)	> 2.00	SID. DEV.		Avg = AVERAGE LABORATORY PERFORMANCE RATING	
LAH	ZN	L/T			
002	NU	L/T			
003	ND	L/T			
004	ND	L/T			
006	ND	L/T			
007	ND	L/T			
008	ND	L/T			
009	ND	L/T			
010	ND	L/T			
011	ND	L/T			
012	ND	L/T			
013	ND	L/T			
014	ND	L/T			
015	L/T	L/T			
016	ND	L/T			
019	NU	L/T			
020	NU	L/T			
022	ND	L/T			
023	ND	L/T			
025	ND	L/T			
026	ND	L/T			
027	NU	L/T			
030	NU	L/T			
032	ND	L/T			
033	ND	L/T			
034	ND	L/T			
035	ND	L/T			
036	ND	L/T			
037	ND	L/T			
039	NU	L/T			
040	NU	L/T			
041	NU	L/T			
042	ND	L/T			
043	ND	L/T			
045	ND	L/T			
049	ND	L/T			
051	NU	L/T			
053	ND	L/T			
054	ND	L/T			
057	ND	L/T			
059	ND	L/T			
065	ND	L/T			

TABLE 4.—OVERALL LABORATORY PERFORMANCE

		SRNS P-5 (PRECIPITATION)											
		NR = NOT RATED					NU = NOT DETERMINED						
		LT = LESS THAN VALUE REPORTED, NOT RATED					N = NUMBER OF CONSTITUENTS LABORATORY DETERMINED						
		AVG = AVERAGE LABORATORY PERFORMANCE RATING											
RATING		SRNS P-5 (PRECIPITATION)											
4 EXCELLENT	0.00	10	0.50	STD.	DEV.	NR	NR	NR	NR	NR	NR	SO4	4
3 (GOOD)	0.51	10	1.00	STD.	DEV.	NU	NU	NU	NU	NU	NU		2
2 (SATISFACTORY)	1.01	10	1.50	STD.	DEV.	LT	LT	LT	LT	LT	LT		3
1 (QUESTIONABLE)	1.51	10	2.00	STD.	DEV.	N	N	N	N	N	N		3
0 (POOR)	> 2.00	STD.	DEV.			AVG	AVG	AVG	AVG	AVG	AVG		4
LAB		NH3-N					PH						
001		NA					SO4						
002		LT											
003		LT											
004		LT											
005		LT											
006		ND											
007		ND											
008		ND											
009		ND											
010		ND											
011		ND											
012		ND											
013		ND											
014		ND											
015		ND											
016		ND											
017		ND											
018		ND											
019		ND											
020		ND											
022		ND											

TABLE 4.--OVERALL LABORATORY PERFORMANCE

RATING	4 (EXCELLENT)	SRWS P-5 (PRECIPITATION)					N = NUMBER OF CONSTITUENTS LABORATORY DETERMINED	AVG. = AVERAGE LABORATORY PERFORMANCE RATING
		0.00	10	0.50	SIU.	DEV.		
3 (GOOD)	0.51	10	1.00	STD.	DEV.	ND	2	3.14
2 (SATISFACTORY)	1.01	10	1.50	STD.	DEV.	LT	6	3.17
1 (QUESTIONABLE)	1.51	10	2.00	STD.	DEV.	N	9	3.55
0 (POOR)	>	2.00	SIU.	DEV.		Avg.	6	1.67
							9	1.00
							6	3.00
							7	3.00
							2	3.50
							5	3.00
							3	3.67
							6	3.00
							3	3.13
							0	2.00
							5	3.50
							10	

TABLE 5.—STANDARD REFERENCE SAMPLE M-66 REPORT FOR ALK(CACU3)

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
002	149	1.5	TITRATION
003	155	2.4	TITRATION
004	134	11.4	TITRATION
006	151	0.2	TITRATION
007	146	3.5	NOT REPUTED
008	143	5.5	TITRATION
009	151	0.2	TITRATION
010	150	0.9	TITRATION
011	312	106.2	REJECT
012	152	0.5	TITRATION
014	165	9.0	REJECT
015	296	95.6	NOT REPUTED
017	135	10.8	TITRATION
019	592	99.6	REJECT
020	152	2.4	NOT REPUTED
022	151	0.2	TITRATION
023	176	16.3	NOT REPUTED
025	145	4.2	TITRATION
026	145	4.2	TITRATION, ELECTROMETRIC, AUTOMATED
027	144	4.6	TITRATION
032	152	0.5	NOT REPUTED
033	162	7.1	TITRATION
035	160	5.7	TITRATION
036	242	59.9	REJECT
037	133	12.1	COLUMIMETRIC
038	151	0.2	TITRATION
039	152	0.5	TITRATION
041	150	0.9	TITRATION
042	145	4.2	COLUMIMETRIC
043	150	0.9	TITRATION
045	151	0.2	TITRATION
047	153	1.1	TITRATION
048	156	3.1	TITRATION
053	153	1.1	TITRATION
054	157	3.8	NOT REPUTED
055	152	0.5	TITRATION
058	150	0.9	TITRATION
059	197	30.2	REJECT
061	152	0.5	TITRATION
062	160	5.7	TITRATION
063	151	0.2	TITRATION
064	146	3.5	TITRATION
066	167	10.4	TITRATION
TOTAL RANGE	133	10	MEAN: 151.5
STANDARD DEVIATION	0.4	312	95 % CONFIDENCE INTERVAL OF MEAN
			151.3 + OR - 2.8

TABLE 3.—STANDARD REFERENCE SAMPLE M-86 REPORT FOR 8

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
003	200	10.7	PLASMA, INDUCTIVELY COUPLED
004	180	19.6	PLASMA, INDUCTIVELY COUPLED
006	320	42.9	COLORIMETRIC
012	200	10.7	PLASMA, DIRECT CURRENT
013	210	6.2	PLASMA, INDUCTIVELY COUPLED
015	210	6.2	COLORIMETRIC
017	320	42.9	COLORIMETRIC
025	220	1.6	COLORIMETRIC
027	205	6.5	PLASMA, INDUCTIVELY COUPLED
036	0	100.0	REJECT
037	300	34.0	SPECTROPHOTOMETRIC
042	250	11.6	PLASMA, INDUCTIVELY COUPLED
045	170	24.1	PLASMA, INDUCTIVELY COUPLED
053	200	10.7	PLASMA, INDUCTIVELY COUPLED
058	150	33.0	COLORIMETRIC

TOTAL RANGE 0 TO 320 MEAN: 223.9
 STANDARD DEVIATION 53.9 95 % CONFIDENCE INTERVAL OF MEAN 223.9 + UR - 31.1

TABLE 5.—STANDARD REFERENCE SAMPLE M-86 REPUBLIC OF BH

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
003	15000	657.4	HELIUM
006	380	30.1	NEUTRON ACTIVATION
037	760	39.7	LULURIMETRIC
047	500	8.1	ION SPECIFIC ELECTRODE
053	480	11.8	ION CHROMATOGRAPHY
059	600	10.3	NOT REPORTED
TOTAL RANGE	580	15000	MEAN: 544.0
STANDARD DEVIATION	143.6	10	95 % CONFIDENCE INTRVL OF MEAN
			544.0 + OR - 176.5

TABLE 5.—STANDARD REFERENCE SAMPLE N-86 REPORT FOR CA

REPORTED CODE	VALUÉ	FCI * FROM MEAN	DEV. FROM MEAN	METHODS
002	71.6	1.5	PLASMA, INDUCTIVELY COUPLED	
003	75.0	6.3	PLASMA, INDUCTIVELY COUPLED	
004	75.0	6.3	PLASMA, INDUCTIVELY COUPLED	
006	750.0	965.4	REJECT	
007	64.0	9.3	ATOMIC ABSORPTION, DIRECT TITRATION	
008	72.5	2.5	ATOMIC ABSORPTION, DIRECT	
010	71.0	0.7	ATOMIC ABSORPTION, DIRECT	
011	66.5	2.9	TITRATION	
012	71.6	1.5	PLASMA, DIRECT CURRENT	
013	74.0	4.9	PLASMA, INDUCTIVELY COUPLED	
014	72.0	2.1	ATOMIC ABSORPTION, DIRECT	
015	70.0	0.8	ATOMIC ABSORPTION, DIRECT	
017	72.5	2.8	TITRATION	
018	72.0	2.1	ATOMIC ABSORPTION, DIRECT	
019	76.0	7.8	ATOMIC ABSORPTION, DIRECT NOT REPUTED	
020	70.0	0.8	TITRATION	
022	69.0	2.2	ATOMIC ABSORPTION, DIRECT	
023	150.0	112.7	REJECT	
025	68.0	3.6	ATOMIC ABSORPTION, DIRECT	
026	64.0	9.3	PLASMA, INDUCTIVELY COUPLED	
027	70.9	0.5	PLASMA, INDUCTIVELY COUPLED NOT REPUTED	
030	71.1	0.8	ATOMIC ABSORPTION, DIRECT	
032	71.5	1.4	ATOMIC ABSORPTION, DIRECT	
033	71.0	0.7	ION CHROMATOGRAPHY	
034	71.6	1.6	ATOMIC ABSORPTION, DIRECT	
035	74.0	4.9	TITRATION	
036	75.0	6.3	ATOMIC ABSORPTION, DIRECT	
037	65.0	7.8	ATOMIC ABSORPTION, DIRECT	
038	61.5	12.8	ATOMIC ABSORPTION, DIRECT	
039	71.0	0.7	ATOMIC ABSORPTION, DIRECT UNHE	
041	72.0	2.1	ATOMIC ABSORPTION, DIRECT	
042	68.2	3.3	ATOMIC ABSORPTION, DIRECT	
043	77.0	9.2	ATOMIC ABSORPTION, DIRECT	
045	74.0	4.9	TITRATION	
047	81.0	14.8	ATOMIC ABSORPTION, DIRECT	
048	72.0	2.1	ATOMIC ABSORPTION, DIRECT PLASMA, INDUCTIVELY COUPLED	
053	72.0	2.1	ATOMIC ABSORPTION, DIRECT NOT REPUTED	
054	69.6	1.3	ATOMIC ABSORPTION, DIRECT	
055	64.0	9.3	ATOMIC ABSORPTION, DIRECT	
057	64.0	9.3	ATOMIC ABSORPTION, DIRECT	
058	70.0	0.8	ATOMIC ABSORPTION, DIRECT	
059	64.0	9.3	TITRATION	
061	66.2	6.1	ATOMIC ABSORPTION, DIRECT NOT REPUTED	
062	179.0	153.8	REJECT	
063	70.0	0.8	TITRATION	
064	69.6	1.3	ATOMIC ABSORPTION, DIRECT	
066	70.4	0.2		
TOTAL RANGE	61.5	10	750.0	95 % CONFINENCE INTERVAL OF MEAN
STANDARD DEVIATION	3.94	3.94		MEAN: 70.53 + OR - 1.20

TABLE 5.—STANDARD REFERENCE SAMPLE N-86 REPORT FOR CL

REPORTED VALUE	PCT. DEV. FROM MEAN	MEASURED
0.02	46.0	COLORIMETRIC TITRATION
0.03	44.0	3.0 TITRATION
0.04	27.0	39.2 REJECT ION CHROMATOGRAPHY
0.06	43.0	3.2 TITRATION
0.07	43.0	3.2 TITRATION
0.08	42.2	5.0 TITRATION
0.09	41.0	7.7 ION SPECIFIC ELECTRODE
0.10	49.0	10.3 ION CHROMATOGRAPHY
0.11	45.5	2.4 TITRATION
0.12	44.0	1.0 TITRATION
0.13	43.0	3.2 COLORIMETRIC TITRATION
0.14	44.0	1.0 TITRATION
0.15	44.0	1.0 TITRATION
0.17	44.0	1.0 TITRATION
0.18	42.5	4.3 NOT REPORTED
0.19	47.0	5.0 NOT REPORTED
0.20	45.0	1.3 TITRATION
0.22	46.0	3.5 TITRATION
0.23	50.0	12.5 NOT REPORTED
0.25	45.0	1.3 COLORIMETRIC TITRATION
0.26	24.0	46.0 REJECT
0.27	44.5	0.2 COLORIMETRIC TITRATION
0.30	43.0	2.6 NOT REPORTED
0.32	31.0	30.2 REJECT ION CHROMATOGRAPHY
0.33	45.0	1.3 ION CHROMATOGRAPHY
0.34	44.0	1.0 COLORIMETRIC TITRATION
0.35	48.0	8.0 NOT REPORTED
0.36	42.0	5.5 COLORIMETRIC TITRATION
0.37	44.0	1.0 36.7 COLORIMETRIC TITRATION
0.38	42.0	17.4 COLORIMETRIC TITRATION
0.59	42.0	5.5 COLORIMETRIC TITRATION
0.40	49.0	3.2 COLORIMETRIC TITRATION
0.41	37.0	16.7 COLORIMETRIC TITRATION
0.42	45.0	1.3 COLORIMETRIC TITRATION
0.43	47.0	5.8 TITRATION
0.45	46.0	3.5 TITRATION
0.47	54.0	21.5 REJECT TITRATION
0.48	44.0	1.0 TITRATION
0.55	46.0	3.5 ION CHROMATOGRAPHY
0.54	46.0	3.5 COLORIMETRIC TITRATION
0.55	31.2	29.6 REJECT NOT REPORTED
0.57	43.0	3.2 COLORIMETRIC TITRATION
0.58	44.0	1.0 48.0 6.0 TITRATION
0.59	48.0	44.2 0.5 TITRATION
0.61	44.0	100.0 REJECT
0.62	0.0	0.0 TITRATION
0.63	48.0	0.0 TITRATION
0.64	49.0	10.3 42.7 3.9 TITRATION
0.66	0.0	MEDIAN: 44.43 95 % CONFIDENCE INTERVAL OF MEAN: 44.43 ± UN - 0.02
TOTAL RANGE	0.0	10 54.0
STANDARD DEVIATION	2.66	

TABLE 5.—STANDARD REFERENCE SAMPLE M-86 REPORT FROM USRD 160

COUNT	REPUTED VALUE	PCT. DEV. FROM MEAN	METHODS
0.03	64.0	9.2	RESIDUE ON EVAPORATION
0.04	63.2	7.8	GRAVIMETRIC
0.06	58.0	1.0	RESIDUE ON EVAPORATION
0.08	54.1	7.7	NOT REPORTED
0.09	50.7	13.5	NOT REPUTED
0.11	59.4	1.4	NOT REPUTED
0.14	54.0	7.9	NOT REPUTED
0.15	60.0	2.4	GRAVIMETRIC
0.17	57.6	1.7	GRAVIMETRIC
0.23	60.5	3.2	NOT REPUTED
0.25	58.0	1.0	GRAVIMETRIC
0.26	60.0	2.4	RESIDUE ON EVAPORATION
0.27	59.5	1.5	GRAVIMETRIC
0.31	58.2	0.7	GRAVIMETRIC
0.36	60.5	3.2	RESIDUE ON EVAPORATION
0.37	60.8	3.7	GRAVIMETRIC
0.43	59.3	1.2	GRAVIMETRIC
0.53	59.2	1.0	RESIDUE ON EVAPORATION
0.54	59.4	1.4	RESIDUE ON EVAPORATION
0.55	76.2	30.0	REJECT
0.59	51.6	12.0	RESIDUE ON EVAPORATION
0.61	61.6	5.1	GRAVIMETRIC
0.62	57.5	2.2	RESIDUE ON EVAPORATION
0.63	57.0	2.7	RESIDUE ON EVAPORATION
0.64	62.6	6.8	RESIDUE ON EVAPORATION

TOTAL RANGE: 50.7
 STANDARD DEVIATION: 33.2
 MEAN: 586.0
 95 % CONFIDENCE INTERVAL OF MEAN: 586.0 ± OR = 14.0

TABLE 5.—STANDARD REFERENCE SAMPLE M-86 REPORT FOR F

CODE	REPORTED VALUE	PCI. DEV. FROM MEAN	METHODS
U03	1.6	17.5	ION SPECIFIC ELECTRODE
U04	1.7	12.4	ION SPECIFIC ELECTRODE
U06	2.4	23.7	ION CHROMATOGRAPHY
U09	1.9	2.1	ION SPECIFIC ELECTRODE
U10	2.0	3.1	ION SPECIFIC ELECTRODE
U11	1.4	27.6	ION SPECIFIC ELECTRODE
U12	1.6	17.5	COLUMNETRIC
U15	1.9	2.1	ION SPECIFIC ELECTRODE
U17	1.5	22.7	ION SPECIFIC ELECTRODE
U25	2.1	8.2	ION SPECIFIC ELECTRODE
U26	1.0	48.5	REJECT
U27	2.0	3.1	ION SPECIFIC ELECTRODE
U32	2.0	3.1	ION CHROMATOGRAPHY
U33	2.3	18.6	ION CHROMATOGRAPHY
U34	1.9	2.1	ION SPECIFIC ELECTRODE
U36	2.0	3.1	NUT REPUTED
U37	4.3	121.6	REJECT
U38	2.2	13.4	ION SPECIFIC ELECTRODE
U40	3.3	70.1	REJECT
U43	2.0	3.1	ION SPECIFIC ELECTRODE
U45	2.2	13.4	ION SPECIFIC ELECTRODE
U47	2.0	3.1	ION SPECIFIC ELECTRODE
U53	1.9	2.1	ION SPECIFIC ELECTRODE
U54	2.0	3.1	ION SPECIFIC ELECTRODE
U57	2.0	3.1	ION SPECIFIC ELECTRODE
U58	2.0	3.1	ION SPECIFIC ELECTRODE
U59	1.9	2.1	NUT REPUTED
U61	2.0	3.1	COLUMNETRIC

TOTAL RANGE 1.0
STANDARD DEVIATION 0.23
MEAN: 1.94
95 % CONFIDENCE INTERVAL OF MEAN 1.94 ± 0.10

TABLE 5.—STANDARD REFERENCE SAMPLE N-66 REPORT FOR I

COUNT	REPORTED VALUE	PCT. FROM MEAN	DEV. FROM MEAN	METHODS
0.03	< 100	0.0	0.0	IGNORED 1 UN SPECIFIC ELECTRUDÉ
0.06	< 5	0.0	0.0	IGNORED NEURON ACTIVATION
0.13	20	0.0	0.0	COLORIMETRIC
0.15	14	0.0	0.0	FILTRATION
0.53	< 50	0.0	0.0	IGNORED COLORIMETRIC
0.59	< 0	0.0	0.0	IGNORED NOT REPORTED
TOTAL RANGE	14	TO 20	MEAN:	17.0
STANDARD DEVIATION		4.24		

TABLE 5.—STANDARD REFERENCE SAMPLE M-86 REPORT FOR K

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
U02	2.0	39.0	REJECT
U03	4.2	8.5	PLASMA, INDUCTIVELY COUPLED
U04	5.6	22.0	PLASMA, INDUCTIVELY COUPLED
U06	4.4	4.1	ATOMIC ABSORPTION, DIRECT
U07	7.8	70.0	ATOMIC ABSORPTION, DIRECT
U08	4.4	4.1	EMISSION, FLAME
U09	4.5	1.9	ATOMIC ABSORPTION, DIRECT
U10	5.0	9.0	ATOMIC ABSORPTION, DIRECT
U11	4.4	4.1	ATOMIC ABSORPTION, DIRECT
U12	5.5	19.9	ATOMIC ABSORPTION, DIRECT
U13	4.7	2.4	ATOMIC ABSORPTION, DIRECT
U14	4.4	4.1	EMISSION, FLAME
U15	4.3	6.3	EMISSION, FLAME
U17	4.6	4.6	EMISSION, FLAME
U18	4.3	6.3	EMISSION, FLAME
U19	5.0	9.0	NOT REPUTED
U20	4.7	2.4	EMISSION, FLAME
U22	4.6	0.2	ATOMIC ABSORPTION, DIRECT
U23	4.9	6.8	NOT REPUTED
U25	5.0	9.0	ATOMIC ABSORPTION, DIRECT
U26	4.4	4.1	EMISSION, FLAME
U27	4.5	1.9	ATOMIC ABSORPTION, DIRECT
U30	4.5	1.9	NOT REPUTED
U32	4.5	1.9	ATOMIC ABSORPTION, DIRECT
U33	4.8	4.6	ATOMIC ABSORPTION, DIRECT
U34	4.9	6.8	ATOMIC ABSORPTION, DIRECT
U35	4.6	0.2	EMISSION, FLAME
U36	5.0	9.0	NOT REPUTED
U37	4.5	4.1	ATOMIC ABSORPTION, DIRECT
U38	4.4	1.9	ATOMIC ABSORPTION, DIRECT
U39	4.1	10.6	ATOMIC ABSORPTION, DIRECT
U41	4.8	4.6	ATOMIC ABSORPTION, DIRECT
U42	4.6	0.2	EMISSION, FLAME
U43	5.1	11.1	ATOMIC ABSORPTION, DIRECT
U45	4.9	6.8	ATOMIC ABSORPTION, DIRECT
U47	4.0	12.8	EMISSION, FLAME
U48	4.2	8.5	ATOMIC ABSORPTION, DIRECT
U53	4.3	6.3	ATOMIC ABSORPTION, DIRECT
U54	4.4	4.1	NOT REPUTED
U57	4.2	8.5	EMISSION, FLAME
U58	4.3	6.3	ATOMIC ABSORPTION, DIRECT
U59	4.6	0.2	ATOMIC ABSORPTION, DIRECT
U61	4.3	6.3	ATOMIC ABSORPTION, DIRECT
U63	5.1	11.1	EMISSION, FLAME
U64	3.9	15.0	ATOMIC ABSORPTION, DIRECT
U66	4.3	6.3	
TOTAL RANGE	2.6	10	MEAN: 4.59
STANDARD DEVIATION	0.37	7.8	% CONFIDENCE INTERVAL OF MEAN: 4.59 + OK - 0.11

TABLE 5.—STANDARD REFERENCE SAMPLE M-86 REPORT FOR MG

REPORTED VALUE	PC1. FRM MEAN	DEV. FRM MEAN	METHODS
002	51.2	12.8	PLASMA, INDUCTIVELY COUPLED
003	29.0	4.9	PLASMA, INDUCTIVELY COUPLED
004	30.0	6.5	PLASMA, INDUCTIVELY COUPLED
006	29.0	4.9	ATOMIC ABSORPTION, DIRECT
007	40.0	73.6	REJECT ILLUMINATION
008	28.3	2.4	ATOMIC ABSORPTION, DIRECT
009	22.0	20.4	ATOMIC ABSORPTION, DIRECT
010	28.0	1.3	ATOMIC ABSORPTION, DIRECT
011	28.7	3.6	ILLUMINATION
012	26.6	3.6	PLASMA, DIRECT CURRENT
013	31.0	12.1	PLASMA, INDUCTIVELY COUPLED
014	28.0	1.3	ATOMIC ABSORPTION, DIRECT
015	25.0	9.6	ATOMIC ABSORPTION, DIRECT
017	27.4	0.9	ILLUMINATION
018	27.7	0.2	ATOMIC ABSORPTION, DIRECT
019	26.0	6.0	NOT REPORTED
020	30.0	8.5	ILLUMINATION
022	27.2	1.6	ATOMIC ABSORPTION, DIRECT
023	100.0	261.7	NOT REPORTED
025	28.0	1.3	ATOMIC ABSORPTION, DIRECT
026	40.0	44.7	REJECT I
027	27.0	2.3	PLASMA, INDUCTIVELY COUPLED
030	29.2	5.6	NOT REPORTED
032	29.3	6.0	ATOMIC ABSORPTION, DIRECT
033	26.6	3.6	ATOMIC ABSORPTION, DIRECT
034	27.3	1.3	ATOMIC ABSORPTION, DIRECT
035	26.0	6.0	ILLUMINATION
036	27.0	2.3	ATOMIC ABSORPTION, DIRECT
037	26.0	6.0	ATOMIC ABSORPTION, DIRECT
038	26.3	4.9	ATOMIC ABSORPTION, DIRECT
039	25.2	6.9	ATOMIC ABSORPTION, DIRECT
041	26.0	6.0	ATOMIC ABSORPTION, DIRECT
042	27.7	0.2	ATOMIC ABSORPTION, DIRECT
043	28.0	1.3	ATOMIC ABSORPTION, DIRECT
045	28.0	1.3	ILLUMINATION
047	25.0	16.6	ATOMIC ABSORPTION, DIRECT
048	28.0	1.3	PLASMA, INDUCTIVELY COUPLED
053	28.0	1.3	ATOMIC ABSORPTION, DIRECT
054	27.0	2.3	NOT REPORTED
055	28.2	2.0	ATOMIC ABSORPTION, DIRECT
057	26.0	6.0	ATOMIC ABSORPTION, DIRECT
058	27.0	2.3	ILLUMINATION
059	29.0	4.9	ATOMIC ABSORPTION, DIRECT
061	26.6	3.6	ATOMIC ABSORPTION, DIRECT
062	111.0	301.5	NOT REPORTED
063	29.0	4.9	ILLUMINATION
064	29.2	5.6	ATOMIC ABSORPTION, DIRECT
066	27.2	1.6	
TOTAL RANGE	22.0	10	111.0
STANDARD DEVIATION	1.60	9.5	CONFIDENCE INTERVAL OF MEAN
			27.65 + OR - 0.49

TABLE 5.— STANDARD REFERENCE SAMPLE M-86 REPORT FOR NA

REPORTED COUNT	REPORTED VALUE	PCT. DEV. FRM MEAN	METHODS
002	78.9	2.4	PLASMA, INDUCTIVELY COUPLED
003	85.0	10.3	PLASMA, INDUCTIVELY COUPLED
004	81.0	5.1	PLASMA, INDUCTIVELY COUPLED
006	77.0	0.1	ATOMIC ABSORPTION, DIRECT EMISSION, FLAME
007	83.0	7.7	ATOMIC ABSORPTION, DIRECT
008	77.5	0.6	ATOMIC ABSORPTION, DIRECT
009	0.6	99.0	REJECT ALUMIC ABSORPTION, DIRECT
010	76.0	1.4	ATOMIC ABSORPTION, DIRECT
011	81.0	5.1	EMISSION, FLAME
012	76.4	0.8	EMISSION, FLAME
013	82.0	6.4	PLASMA, INDUCTIVELY COUPLED
014	77.0	0.1	NOT REPORTED
015	76.0	1.4	EMISSION, FLAME
017	79.5	3.2	EMISSION, FLAME
01b	52.1	32.4	ALUMIC ABSORPTION, DIRECT
019	75.0	5.3	NOT REPORTED
020	81.0	5.1	EMISSION, FLAME
022	75.0	2.7	ATOMIC ABSORPTION, DIRECT
023	69.0	10.4	NOT REPORTED
025	78.0	1.2	ATOMIC ABSORPTION, DIRECT
026	78.0	1.2	PLASMA, INDUCTIVELY COUPLED
027	77.8	1.0	NOT REPORTED
030	77.1	0.1	ATOMIC ABSORPTION, DIRECT
032	76.0	1.4	EMISSION, FLAME
033	89.0	3.6	ATOMIC ABSORPTION, DIRECT
034	69.4	9.9	ATOMIC ABSORPTION, DIRECT
035	80.0	3.8	EMISSION, FLAME
036	76.0	1.4	EMISSION, FLAME
037	75.0	2.7	ATOMIC ABSORPTION, DIRECT
038	74.9	2.6	ATOMIC ABSORPTION, DIRECT
039	79.0	2.5	ATOMIC ABSORPTION, DIRECT
041	70.0	9.1	ATOMIC ABSORPTION, DIRECT
042	78.4	2.3	EMISSION, FLAME
043	78.0	1.2	ATOMIC ABSORPTION, DIRECT
045	74.0	4.0	ATOMIC ABSORPTION, DIRECT
047	68.0	11.7	EMISSION, FLAME
048	80.0	1.2	ATOMIC ABSORPTION, DIRECT
053	82.0	3.6	PLASMA, INDUCTIVELY COUPLED
054	77.0	6.4	NOT REPORTED
057	76.0	0.1	EMISSION, FLAME
058	78.0	1.2	ATOMIC ABSORPTION, DIRECT
059	80.0	3.6	ATOMIC ABSORPTION, DIRECT
061	75.0	2.7	ATOMIC ABSORPTION, DIRECT
063	71.0	7.9	NOT REPORTED
064	76.5	0.7	EMISSION, FLAME
066	74.4	3.4	ATOMIC ABSORPTION, DIRECT

TOTAL RANGE
STANDARD DEVIATIONMEAN: 77.05
95 % CONFIDENCE INTERVAL OF MEAN 77.05 ± 0K = 1.13

TABLE 5.—STANDARD REFERENCE SAMPLE M-86 REPORT FOR NU2-N

CODE	REPORTED VALUE	PCT. FROM MEAN	STD. DEV.	METHODS
U03	0.07	63.1	COLORIMETRIC	
0.04	0.02	53.4	COLORIMETRIC	
U06	0.02	53.4	COLORIMETRIC	
U10	0.15	249.5	COLORIMETRIC	
U11	0.01	76.7	COLORIMETRIC	
U12	0.01	76.7	COLORIMETRIC	
U14	<	IGNORED	COLORIMETRIC	
U15	0.02	53.4	COLORIMETRIC	
U17	0.18	319.4	COLORIMETRIC	
U18	0.03	30.1	COLORIMETRIC	
U20	0.01	76.7	COLORIMETRIC	
U22	0.01	76.7	COLORIMETRIC	
U25	0.04	6.8	COLORIMETRIC	
U26	0.02	53.4	COLORIMETRIC	
U27	0.01	76.7	COLORIMETRIC	
U35	0.03	30.1	COLORIMETRIC	
U37	<	IGNORED	COLORIMETRIC	
U38	0.01	76.7	COLORIMETRIC	
U39	0.16	272.8	COLORIMETRIC	
U40	0.01	76.7	COLORIMETRIC	
U43	0.33	668.9	REJECT	
U48	0.25	482.5	REJECT	
U53	0.02	53.4	10N CHROMATOGRAPHY	
U55	0.45	948.6	REJECT	
U57	0.11	156.3	NOT REPORTED	
U59	0.00	100.0	COLORIMETRIC	
U61	0.02	53.4	COLORIMETRIC	
U62	0.06	39.6	10N SPECIFIC ELECTRUE	
U64	0.01	76.7	COLORIMETRIC	

TOTAL RANGE 0.00 TO 0.45 MEANS 0.043 95 % CONFIDENCE INTERVAL OF MEAN 0.043 + OR - 0.022

TABLE 5.—STANDARD REFERENCE SAMPLE N-86 REPORT FOR NBS-N

COUNT	REPORTED VALUE	PCI FROM MEAN	DEV. FROM MEAN	METHODS
003	4.30	11.3		COLORIMETRIC
004	3.78	2.1		COLORIMETRIC
006	5.93	1.7		COLORIMETRIC
007	2.80	27.5		NIN REPUTED
008	3.85	0.3		COLORIMETRIC
009	7.31	0.92	REJECT	COLORIMETRIC
010	4.00	3.6		1IN CHROMATOGRAPHY
011	3.88	0.4		COLORIMETRIC
012	4.05	4.9		COLORIMETRIC
013	3.70	4.2		COLORIMETRIC
014	3.61	6.5		COLORIMETRIC
015	4.40	13.9		COLORIMETRIC
017	4.69	21.2		COLORIMETRIC
018	1.76	54.4	REJECT	COLORIMETRIC
020	4.40	13.9		COLORIMETRIC
022	3.62	6.3		COLORIMETRIC
025	4.10	6.1		COLORIMETRIC
026	0.91	76.4	REJECT	COLORIMETRIC
027	4.09	5.9		COLORIMETRIC
032	16.80	334.9	REJECT	1IN CHROMATOGRAPHY
033	5.70	4.2		COLORIMETRIC
034	3.60	6.8		COLORIMETRIC
035	4.00	3.6		COLORIMETRIC
036	3.15	18.4		COLORIMETRIC
057	4.40	13.9		COLORIMETRIC
058	3.60	6.6		COLORIMETRIC
039	2.00	46.7	REJECT	COLORIMETRIC
040	4.04	4.6		COLORIMETRIC
041	3.20	14.8		COLORIMETRIC
043	5.76	2.7		COLORIMETRIC
045	17.00	340.1	REJECT	1IN SPECIFIC ELECTRODE
047	13.00	236.6	REJECT	1IN SPECIFIC ELECTRODE
048	3.85	0.3		COLORIMETRIC
053	4.11	6.4		1IN CHROMATOGRAPHY
054	3.30	14.6		1IN SPECIFIC ELECTRODE
057	3.81	1.4		COLORIMETRIC
058	3.65	6.0		COLORIMETRIC
059	3.90	1.0		COLORIMETRIC
061	3.95	2.3		COLORIMETRIC
062	6.80	16.0	REJECT	1IN SPECIFIC ELECTRODE
064	4.10	6.1		COLORIMETRIC
066	3.95	2.3		COLORIMETRIC

TOTAL RANGE 0.91
STANDARD DEVIATION 0.379
MEAN: 3.663
95 % CONFIDENCE INTERVAL OF MEAN 3.863 + OR - 0.132

TABLE 5.—STANDARD REFERENCE SAMPLE M-86 REPORT FOR P, TOTAL

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
002	0.46	4.2	COLORIMETRIC
003	0.48	0.0	COLORIMETRIC
004	4.50	83.5	REJECT
006	0.50	4.2	COLORIMETRIC
009	4.99	939.6	REJECT
010	0.90	67.5	NUT REPUTED
011	0.34	29.2	COLORIMETRIC
012	0.64	33.3	COLORIMETRIC
013	0.52	8.3	PLASMA, INDUCTIVELY COUPLED
014	0.45	6.2	COLORIMETRIC
015	0.47	2.1	COLORIMETRIC
017	0.57	22.9	COLORIMETRIC
018	0.46	4.2	COLORIMETRIC
020	0.48	0.0	COLORIMETRIC
022	0.54	12.5	COLORIMETRIC
023	0.46	4.2	NUT REPUTED
025	0.49	2.1	COLORIMETRIC
026	0.45	6.2	COLORIMETRIC
027	0.53	10.4	COLORIMETRIC
034	0.52	6.3	COLORIMETRIC
035	0.52	8.3	COLORIMETRIC
036	0.32	33.3	NUT REPUTED
037	0.55	14.6	COLORIMETRIC
038	0.47	2.1	COLORIMETRIC
039	0.46	0.0	COLORIMETRIC
041	0.47	2.1	COLORIMETRIC
042	0.48	0.0	COLORIMETRIC
043	0.47	2.1	COLORIMETRIC
045	0.60	25.0	PLASMA, INDUCTIVELY COUPLED
048	0.48	0.0	COLORIMETRIC
053	0.52	33.3	COLORIMETRIC
054	0.52	8.3	OTHER
057	0.52	8.3	COLORIMETRIC
058	0.49	2.1	COLORIMETRIC
059	0.55	14.6	COLORIMETRIC
064	0.50	4.2	COLORIMETRIC
066	0.42	12.5	COLORIMETRIC
TOTAL RANGE	0.52	4.99	MEAN: 0.480
STANDARD DEVIATION	0.068	95 % CONFIDENCE INTERVAL OF MEAN	0.480 + OR - 0.024

TABLE 5.—STANDARD REFERENCE SAMPLE M-86 REPORT FOR PH

REPORTED VALUE	CODE	PCT. DEV.	FROM MEAN	METHODS
0.02	U02	7.8	0.8	ELECTRUMETRIC
0.03	U03	7.8	0.8	NUT REPOUNDED
0.04	U04	7.6	3.3	ELECTRUMETRIC
0.06	U06	7.8	0.8	ELECTRUMETRIC
0.07	U07	7.6	3.3	ELECTRUMETRIC
0.08	U08	7.3	3.3	ELECTRUMETRIC
0.10	U10	6.1	7.1	ELECTRUMETRIC
0.11	U11	6.0	3.0	ELECTRUMETRIC
0.12	U12	6.0	1.8	NUT REPOUNDED
0.13	U13	6.1	3.3	ELECTRUMETRIC
0.14	U14	7.5	4.6	NUT REPOUNDED
0.15	U15	7.6	3.3	ELECTRUMETRIC
0.17	U17	7.8	0.8	ELECTRUMETRIC
0.18	U18	7.8	0.8	NUT REPOUNDED
0.19	U19	7.9	0.5	ELECTRUMETRIC
0.20	U20	7.8	0.8	ELECTRUMETRIC
0.22	U22	8.0	1.8	NUT REPOUNDED
0.23	U23	8.0	1.8	ELECTRUMETRIC
0.25	U25	7.8	0.8	NUT REPOUNDED
0.26	U26	7.8	0.8	ELECTRUMETRIC
0.27	U27	7.7	2.0	ELECTRUMETRIC
0.30	U30	7.9	0.5	NUT REPOUNDED
0.32	U32	8.0	1.8	NUT REPOUNDED
0.33	U33	8.5	8.1	ELECTRUMETRIC
0.34	U34	7.9	0.5	ELECTRUMETRIC
0.35	U35	7.8	0.8	ELECTRUMETRIC
0.36	U36	8.2	4.3	ELECTRUMETRIC
0.37	U37	7.7	2.0	ELECTRUMETRIC
0.38	U38	7.7	2.0	ELECTRUMETRIC
0.39	U39	7.9	0.5	ELECTRUMETRIC
0.41	U41	8.1	3.0	ELECTRUMETRIC
0.42	U42	7.7	2.0	ELECTRUMETRIC
0.43	U43	7.7	2.0	ELECTRUMETRIC
0.45	U45	8.1	3.0	NUT REPOUNDED
0.48	U48	7.8	0.8	ELECTRUMETRIC
0.53	U53	7.7	2.0	ELECTRUMETRIC
0.54	U54	7.8	0.8	ELECTRUMETRIC
0.55	U55	8.2	4.3	NUT REPOUNDED
0.57	U57	8.0	1.8	ELECTRUMETRIC
0.58	U58	8.3	5.6	ELECTRUMETRIC
0.59	U59	7.5	4.6	ELECTRUMETRIC
0.61	U61	8.0	1.8	ELECTRUMETRIC
0.62	U62	8.1	3.0	ELECTRUMETRIC
0.63	U63	7.6	0.8	ELECTRUMETRIC
0.64	U64	8.0	1.8	ELECTRUMETRIC
0.66	U66	7.8	0.8	ELECTRUMETRIC
TOTAL RANGE		7.3	10	8.5
STANDARD DEVIATION		0.23	9.5 % CONFIDENCE INTERVAL OF MEAN	7.86 ± UN = 0.07

TABLE 5.--STANDARD REFERENCE SAMPLE M-86 REPORT FOR SIU2

REPURIFIED CUT	REPURIFIED VALUE	PCT. DEV. FROM MEAN	METHODS
002	13.2	10.6	COLORIMETRIC
003	14.4	20.8	PLASMA, INDUCTIVELY COUPLED
004	6.7	43.8	PLASMA, INDUCTIVELY COUPLED
006	12.6	5.7	COLORIMETRIC
007	6.0	49.7	COLORIMETRIC
008	14.2	61.1	COLORIMETRIC
010	0.5	97.5	COLORIMETRIC
011	5.3	55.5	COLORIMETRIC
012	6.5	45.5	PLASMA, DIRECT CURRENT
013	12.3	3.2	PLASMA, INDUCTIVELY COUPLED
014	13.5	13.3	COLORIMETRIC
015	12.3	3.2	COLORIMETRIC
018	6.6	44.6	COLORIMETRIC
020	13.0	9.1	COLORIMETRIC
022	12.0	0.7	COLORIMETRIC
025	14.0	17.5	COLORIMETRIC
026	13.0	9.1	COLORIMETRIC
027	6.2	46.0	PLASMA, INDUCTIVELY COUPLED
030	11.8	1.0	NOT REPORTED
032	7.5	37.1	COLORIMETRIC
033	13.0	9.1	ALUMIC ABSORPTION, DIRECT
034	10.0	16.1	COLORIMETRIC
035	15.5	30.1	COLORIMETRIC
036	14.0	17.5	COLORIMETRIC
037	16.6	39.3	COLORIMETRIC
038	12.2	2.4	COLORIMETRIC
040	13.2	10.6	COLORIMETRIC
041	13.1	9.9	COLORIMETRIC
042	14.2	19.1	COLORIMETRIC
043	13.6	15.6	COLORIMETRIC
045	9.0	24.5	PLASMA, INDUCTIVELY COUPLED
048	14.1	16.3	COLORIMETRIC
053	12.3	3.2	PLASMA, INDUCTIVELY COUPLED
054	14.0	17.5	ALUMIC ABSORPTION, DIRECT
057	13.5	13.3	COLORIMETRIC
058	13.4	12.4	COLORIMETRIC
059	15.0	25.9	COLORIMETRIC
061	6.5	45.5	COLORIMETRIC
062	11.6	2.7	NOT REPORTED
066	13.7	15.0	COLORIMETRIC
TOTAL RANGE	0.5	19.2	MEAN: 11.92
STANDARD DEVIATION	3.27	95 % CONFIDENCE INTERVAL OF MEAN	11.92 ± 0.4 - 1.06

TABLE 5.—STANDARD REFERENCE SAMPLE H-86 REPORT FOR S04

REPORTED CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
002	220	0.2	CULORIMETRIC
003	230	4.4	PLASMA, INDUCTIVELY COUPLED
004	190	13.8	TURBIDIMETRIC
006	226	3.4	ION CHROMATOGRAPHY
007	198	10.2	TISSRATION
008	234	6.2	GRAVIMETRIC
009	227	3.0	GRAVIMETRIC
010	206	6.5	ION CHROMATOGRAPHY
011	223	1.2	GRAVIMETRIC
012	226	2.5	TURBIDIMETRIC
013	232	5.3	PLASMA, INDUCTIVELY COUPLED
014	216	2.0	ATOMIC ABSORPTION, DIRECT
015	219	0.6	GRAVIMETRIC
017	230	4.4	TURBIDIMETRIC
019	222	0.7	NOT REPUTED
020	254	15.2	GRAVIMETRIC
022	226	3.4	TISSRATION
023	180	18.3	NOT REPUTED
025	219	0.6	CULORIMETRIC
026	200	9.3	ION CHROMATOGRAPHY
027	305	38.4	REJECT
030	214	2.9	NOT REPUTED
032	204	7.4	ION CHROMATOGRAPHY
033	224	1.6	TISSRATION
034	260	18.0	CULORIMETRIC
035	198	10.2	GRAVIMETRIC
036	229	3.9	NOT REPUTED
037	213	3.4	CULORIMETRIC
038	226	2.5	TURBIDIMETRIC
039	233	5.7	CULORIMETRIC
040	217	1.5	CULORIMETRIC
041	200	9.3	ION CHROMATOGRAPHY
042	233	5.7	TURBIDIMETRIC
043	221	0.3	GRAVIMETRIC
045	194	12.0	GRAVIMETRIC
047	210	4.7	TURBIDIMETRIC
048	236	6.0	TISSRATION
053	220	0.2	ION CHROMATOGRAPHY
054	239	8.4	TURBIDIMETRIC
055	190	10.4	REJECT
057	225	2.1	CULORIMETRIC
058	229	3.9	CULORIMETRIC
059	205	7.0	CULORIMETRIC
061	222	0.7	ION CHROMATOGRAPHY
062	220	0.2	NOT REPUTED
063	51	76.9	REJECT
064	234	6.2	TURBIDIMETRIC
066	228	3.4	CULORIMETRIC

TOTAL RANGE: 51
STANDARD DEVIATION: 15.7MEAN: 220.4
95 % LIMIT: 191.0 - 220.4 + 0.44.7
220.4 + 0.4 -

TABLE 5.—STANDARD REFERENCE SAMPLE M-86 REPORT FOR SP. COND.

REPORTED CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
002	893	3.9	DIRECT READING INSTRUMENT
003	935	6.7	NOT REPORTED
004	910	5.6	DIRECT READING INSTRUMENT
006	900	4.7	DIRECT READING INSTRUMENT
007	900	4.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER
008	790	6.1	DIRECT READING INSTRUMENT
009	103	66.0	REJECT NOT REPORTED
010	900	4.7	DIRECT READING INSTRUMENT
011	780	9.3	NOT REPORTED
012	875	1.6	DIRECT READING INSTRUMENT
013	569	33.6	REJECT NOT REPORTED
014	604	6.5	DIRECT READING INSTRUMENT
015	895	4.1	NOT REPORTED
017	879	2.2	DIRECT READING INSTRUMENT
018	540	37.2	DIRECT READING INSTRUMENT
019	900	4.7	NOT REPORTED
020	735	14.5	DIRECT READING INSTRUMENT
022	881	2.5	NOT REPORTED
023	875	1.6	DIRECT READING INSTRUMENT
025	903	5.0	DIRECT READING INSTRUMENT
026	879	2.2	DIRECT READING INSTRUMENT
027	691	3.6	DIRECT READING INSTRUMENT
030	692	3.7	NOT REPORTED
032	842	2.1	DIRECT READING INSTRUMENT
033	765	11.0	DIRECT READING INSTRUMENT
034	887	3.2	DIRECT READING INSTRUMENT
035	879	2.2	DIRECT READING INSTRUMENT
036	900	4.7	DIRECT READING INSTRUMENT
037	707	17.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER
038	889	3.4	DIRECT READING INSTRUMENT
039	880	2.3	DIRECT READING INSTRUMENT
041	671	22.0	REJECT NOT REPORTED
042	901	4.6	DIRECT READING INSTRUMENT
043	632	3.2	DIRECT READING INSTRUMENT
045	760	9.3	NOT REPORTED
048	811	5.7	DIRECT READING INSTRUMENT
053	920	7.0	DIRECT READING INSTRUMENT
054	894	4.0	DIRECT READING INSTRUMENT
055	860	0.0	NOT REPORTED
057	790	6.1	DIRECT READING INSTRUMENT
058	894	4.0	DIRECT READING INSTRUMENT
059	820	4.6	DIRECT READING INSTRUMENT
061	634	3.0	DIRECT READING INSTRUMENT
062	1438	67.2	REJECT NOT REPORTED
063	869	1.1	DIRECT READING INSTRUMENT
064	840	2.3	DIRECT READING INSTRUMENT
066	901	4.8	DIRECT READING INSTRUMENT
TOTAL RANGE	103	14.8	MEANS: 859.8
STANDARD DEVIATION	53.1	9.5	SUM OF INFLUENCE FACTOR + UK = 459.8 + 0K = 16.5

TABLE 5.—STANDARD REFERENCE SAMPLE M-86 REPORT FOR SR

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
003	780	1.9	PLASMA, INDUCTIVELY COUPLED
006	730	4.6	ATOMIC ABSORPTION, DIRECT
008	780	1.9	ATOMIC ABSORPTION, DIRECT
009	940	22.8	ATOMIC ABSORPTION, DIRECT
010	800	4.5	PLASMA, INDUCTIVELY COUPLED
012	790	3.2	PLASMA, DIRECT CURRENT
013	760	0.7	PLASMA, INDUCTIVELY COUPLED
014	780	1.9	ATOMIC ABSORPTION, DIRECT
015	990	29.4	ATOMIC ABSORPTION, DIRECT
022	670	12.5	ATOMIC ABSORPTION, DIRECT
025	590	22.9	ATOMIC ABSORPTION, DIRECT
027	755	4.0	PLASMA, INDUCTIVELY COUPLED
039	750	2.0	ATOMIC ABSORPTION, DIRECT
042	720	5.9	PLASMA, INDUCTIVELY COUPLED
045	700	8.5	PLASMA, INDUCTIVELY COUPLED
054	730	4.6	PLASMA, INDUCTIVELY COUPLED

TOTAL RANGE 590 MEAN: 765.3
 STANDARD DEVIATION 94.3 95 % CONFIDENCE INTERVAL OF MEAN 765.3 + OR - 50.2

TABLE 5.—STANDARD REFERENCE SAMPLE M-86 REPORT FOR V

CODE	REFURBED VALUE	PCT. DEV. FROM MEAN	METHODS	REFERENCES
003	6	44.3	REJECT	PLASMA, INDUCTIVELY COUPLED
004	305	729.9	REJECT	PLASMA, INDUCTIVELY COUPLED
006	19	76.3	REJECT	NEUTRIN ACTIVATION
010	<	***	IGNORED	PLASMA,
012	12	11.3	DIRECT CURRENT	PLASMA,
013	11	2.1	INDUCTIVELY COUPLED	PLASMA,
014	12	11.3	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
025	10	7.2	IGNITED ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
026	<	***	IGNITED ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
027	10	7.2	PLASMA, INDUCTIVELY COUPLED	PLASMA, INDUCTIVELY COUPLED
032	11	2.1	ATOMIC ABSORPTION, FLAMELESS	ATOMIC ABSORPTION, FLAMELESS
042	9	16.5	PLASMA, INDUCTIVELY COUPLED	PLASMA, INDUCTIVELY COUPLED
045	12	11.3	PLASMA, INDUCTIVELY COUPLED	PLASMA, INDUCTIVELY COUPLED
053	10	7.2		

TOTAL RANGE 6 STANDARD DEVIATION 1.1 MEAN: 10.8
 95 % CONFIDENCE INTERVAL OF MEAN 10.8 + OR - 0.6

TABLE 6.—STATISTICS BY METHOD FOR SAMPLE: M-86

DETERMINATION: ALK (CACUS)			
METHOD	MEAN	STD DEV	N
TITRATION	151.5	7.3	30
***** OVER ALL *****	151.3	6.4	36
DETERMINATION: B			
METHOD	MEAN	STD DEV	N
COLORIMETRIC	253.3	70.3	6
PLASMA, INDUCTIVELY COUPLED	202.1	25.5	7
***** OVER ALL *****	223.9	53.9	14
DETERMINATION: BR			
METHOD	MEAN	STD DEV	N
***** OVER ALL *****	344.0	143.6	5
DETERMINATION: CA			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT	69.62	3.46	20
PLASMA, INDUCTIVELY COUPLED	71.79	3.01	7
FLUORIM.	71.15	4.94	11
***** OVER ALL *****	70.53	3.94	44
DETERMINATION: CL			
METHOD	MEAN	STD DEV	N
COLORIMETRIC	45.13	2.70	16
ION CHROMATOGRAPHY	45.75	2.50	4
FLUORIM.	45.23	2.06	16
***** OVER ALL *****	44.43	2.68	43

TABLE 6.--STATISTICS BY METHOD FOR SAMPLE: H-86

DETERMINATION: DSRD 180			
METHOD	MEAN	STD DEV	N
GRAVIMETRIC	598.0	18.3	9
RESIDUE ON EVAPORATION	589.6	34.0	10
***** OVER ALL *****	586.0	33.2	24
DETERMINATION: F			
METHOD	MEAN	STD DEV	N
ION CHROMATOGRAPHY	2.23	0.21	3
ION SPECIFIC ELECTRODE	1.91	0.22	18
***** OVER ALL *****	1.94	0.23	25
DETERMINATION: I			
METHOD	MEAN	STD DEV	N
***** INSUFFICIENT DATA *****			
DETERMINATION: K			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT EMISSION, FLAME	4.61	0.35	23
***** OVER ALL *****	4.47	0.40	13
4.59	0.37	44	
DETERMINATION: MG			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT PLASMA, INDUCTIVELY COUPLED TITRATION	27.19	1.11	23
***** OVER ALL *****	29.37	1.68	6
27.81	2.15	9	
27.65	1.60	43	
DETERMINATION: NA			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT EMISSION, FLAME	75.96	2.94	18
PLASMA, INDUCTIVELY COUPLED	77.94	3.59	14
***** OVER ALL *****	61.12	2.55	6
77.05	3.73	44	

TABLE 6.—STATISTICS BY METHOD FOR SAMPLE: H-66

DETERMINATION: NO2-N			
METHOD	MEAN	STD DEV	N
COLORIMETRIC	0.043	0.055	22
***** OVER ALL *****	0.043	0.053	24
 DETERMINATION: NO3-N			
METHOD	MEAN	STD DEV	N
COLORIMETRIC	3.411	0.330	29
ION CHROMATOGRAPHY	3.937	0.212	3
***** OVER ALL *****	3.063	0.379	34
 DETERMINATION: P, TOTAL			
METHOD	MEAN	STD DEV	N
COLORIMETRIC	0.479	0.063	29
***** OVER ALL *****	0.460	0.066	34
 DETERMINATION: PH			
METHOD	MEAN	STD DEV	N
ELECTRODE INIC	7.03	0.24	35
***** OVER ALL *****	7.06	0.23	46
 DETERMINATION: S102			
METHOD	MEAN	STD DEV	N
COLORIMETRIC	12.39	3.26	26
PLASMA, INFRAREDLY COUPLED	10.15	3.35	6
***** OVER ALL *****	11.92	3.27	39
 DETERMINATION: S04			
METHOD	MEAN	STD DEV	N
COLORIMETRIC	225.0	14.1	11
GRAVIMETRIC	221.3	19.1	6
ION CHROMATOGRAPHY	214.9	11.2	7
TITRATION	221.3	20.0	3
***** OVER ALL *****	223.1	17.3	7
***** OVER ALL *****	220.4	15.7	45

TABLE 6.—STATISTICS BY METHOD FOR SAMPLE: M-86

DETERMINATION: SP. COND.			
METHOD	MEAN	STD DEV	N
DIRECT READING INSTRUMENT	861.6	48.9	29
***** OVER ALL *****	859.6	53.1	42

DETERMINATION: SR			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT	778.0	131.6	8
PLASMA, INDUCTIVELY COUPLED	746.4	35.2	7
***** OVER ALL *****	765.3	94.3	16

DETERMINATION: V			
METHOD	MEAN	STD DEV	N
PLASMA, INDUCTIVELY COUPLED	10.4	1.1	5
***** OVER ALL *****	10.6	1.1	9

TABLE 7.—STANDARD REFERENCE SAMPLE T-67 REPORT FOR ACIDOPACO3

CUDt	REPUNTED VALUE	PCT. DEV. FROM MEAN	METHODS
0.03	393	0.4	ITRATION
0.04	376	4.7	ITRATION
0.06	411	4.1	ITRATION
0.07	369	6.5	NUT REPORTED
0.08	770	95.1	ITRATION
0.09	763	98.4	REJECT
0.10	380	3.7	ITRATION
0.12	388	1.7	ITRATION
0.14	406	2.9	ITRATION
0.23	394	0.2	NUT REPORTED
0.26	400	1.4	OTHER
0.27	426	6.0	ITRATION
0.33	415	5.2	ITRATION
0.43	377	4.5	ITRATION
0.49	411	4.1	ITRATION
0.51	383	2.9	ITRATION
0.54	395	0.1	ITRATION
0.54	390	1.2	NUT REPORTED
0.59	73	81.5	REJECT
			NUT REPORTED

TOTAL RANGE 73
STANDARD DEVIATION 16.0
MEAN: 394.6
95 X CONFIDENCE INTERVAL OF MEAN 394.6 + OR - 8.5

TABLE 7.—STANDARD REFERENCE SAMPLE 1-87 REPORT FOR AG

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
U03	<	5	*** IGNORED PLASMA, INDUCTIVELY COUPLED
U04	2	5.3	*** IGNORED ATOMIC ABSORPTION, DIRECT
U06	<	0	*** IGNORED ATOMIC ABSORPTION, FLAMELESS
U08	1	47.4	*** IGNORED ATOMIC ABSORPTION, DIRECT
U09	2	5.3	*** IGNORED ATOMIC ABSORPTION, DIRECT
U10	<	3	*** IGNORED ATOMIC ABSORPTION, DIRECT
U11	6	215.8	*** IGNORED ATOMIC ABSORPTION, DIRECT
U12	<	10	*** IGNORED NOT REPUTED
U13	<	2	*** IGNORED PLASMA, INDUCTIVELY COUPLED
U14	0	100.0	*** ATOMIC ABSORPTION, DIRECT
U15	<	1	*** IGNORED ATOMIC ABSORPTION, DIRECT
U25	<	1	*** IGNORED ATOMIC ABSORPTION, DIRECT
U26	<	30	*** IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)
U27	1	47.4	*** IGNORED ATOMIC ABSORPTION, FLAMELESS
U36	3	57.9	*** ATOMIC ABSORPTION, DIRECT
U37	3	57.9	*** ATOMIC ABSORPTION, DIRECT
U45	<	1	*** IGNORED PLASMA, INDUCTIVELY COUPLED
U49	0	100.0	*** NOT REPUTED
U53	<	50	*** IGNORED ATOMIC ABSORPTION, DIRECT
U59	<	1	47.4 IGNORED ATOMIC ABSORPTION, FLAMELESS

TOTAL RANGE 0 10
 STANDARD DEVIATION 0 1.6
 MEAN: 1.9
 95 X CONFIDENCE INTERVAL OF MEAN 1.9 + OR - 1.3

TABLE 7.--STANDARD REFERENCE SAMPLE 1-87 REPORT FOR AL

REPORTED CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
U03	<	50	IGNITED PLASMA, INDUCTIVELY COUPLED REJECT PLASMA, INDUCTIVELY COUPLED
U04	990	145.9	IGNITED ATOMIC ABSORPTION, DIRECT
U06	<	100	ATOMIC ABSORPTION, DIRECT
U08	33	8.2	ATOMIC ABSORPTION, DIRECT
U12	90	195.1	PLASMA, DIRECT CURRENT
U13	3	90.2	PLASMA, INDUCTIVELY COUPLED
U14	5	83.6	ATOMIC ABSORPTION, DIRECT
U15	12	60.7	ATOMIC ABSORPTION, DIRECT
U22	4	86.9	NOT REPUTED
U25	<	200	IGNITED ATOMIC ABSORPTION, DIRECT
U26	<	600	IGNITED ATOMIC ABSORPTION, CHELATION EXTRACTION, NITROUS OXIDE, MANUAL
U27	50	63.9	PLASMA, INDUCTIVELY COUPLED
U30	0	100.0	NOT REPUTED
U37	510	572.1	REJECT NOT REPUTED
U39	<	150	IGNITED ATOMIC ABSORPTION, DIRECT
U40	5	83.6	ATOMIC ABSORPTION, DIRECT
U42	<	50	IGNITED PLASMA, INDUCTIVELY COUPLED
U43	70	129.5	NOT REPUTED
U45	10	67.2	PLASMA, INDUCTIVELY COUPLED
U49	64	175.4	NOT REPUTED
U53	<	10	IGNITED NOT REPUTED
TOTAL RANGE	0	990	MEAN: 30.5
STANDARD DEVIATION	0	34.1	95 % CONFIDENCE INTERVAL OF MEAN 30.5 + OR - 21.7

TABLE 7.—STANDARD REFERENCE SAMPLE I-87 REPORT FOR AS

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS	
			ATOMIC ABSORPTION, FLAMELESS	ATOMIC ABSORPTION, HYDRIDE
002	1	50.0	IGNORED	ATOMIC ABSORPTION, HYDRIDE
003	<	***	IGNITED	ATOMIC ABSORPTION, HYDRIDE
004	<	10	***	IGNITED
006	2	0.0	REJECT	ATOMIC ABSORPTION, HYDRIDE
009	13	550.0	***	IGNORED
011	<	2	0.0	ATOMIC ABSORPTION, FLAMELESS
U14	2	0.0	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, FLAMELESS
015	2	0.0	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, HYDRIDE
025	2	0.0	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, HYDRIDE
026	2	0.0	ATOMIC ABSORPTION, HYDRIDE	ATOMIC ABSORPTION, HYDRIDE
027	2	0.0	ATOMIC ABSORPTION, HYDRIDE	ATOMIC ABSORPTION, HYDRIDE
033	4	100.0	REJECT	PLASMA, INDUCTIVELY COUPLED
040	2	0.0	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
042	<	20	***	IGNORED PLASMA, INDUCTIVELY COUPLED
U43	<	20	***	IGNORED NUT REPORTED
U45	<	2	***	IGNORED ATOMIC ABSORPTION, HYDRIDE
049	18	800.0	REJECT	NUT REPORTED
053	3	50.0	ATOMIC ABSORPTION, HYDRIDE	ATOMIC ABSORPTION, HYDRIDE
057	2	0.0	ATOMIC ABSORPTION, HYDRIDE	ATOMIC ABSORPTION, HYDRIDE

TOTAL RANGE 1 STANDARD DEVIATION 0.5 MEAN: 2.0
 16 95 % CONFIDENCE INTRVL OF MEAN 2.0 + OR - 0.3

TABLE 7.—STANDARD REFERENCE SAMPLE 1-87 REPORT FOR BA

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
003	72	9.5	PLASMA, INDUCTIVELY COUPLED
006	<	**†	IGNORED ATOMIC ABSORPTION, DIRECT
010	70	6.4	PLASMA, INDUCTIVELY COUPLED
012	58	11.6	PLASMA, DIRECT CURRENT
013	58	11.6	PLASMA, INDUCTIVELY COUPLED
014	50	24.0	ATOMIC ABSORPTION, DIRECT
015	56	14.9	ATOMIC ABSORPTION, FLAMELESS
025	50	24.0	ATOMIC ABSORPTION, DIRECT
026	100	52.0	ATOMIC ABSORPTION, DIRECT
027	56	14.9	PLASMA, INDUCTIVELY COUPLED
030	<	**†	IGNORED NOT REPORTED
042	60	6.6	PLASMA, INDUCTIVELY COUPLED
045	60	6.6	PLASMA, INDUCTIVELY COUPLED
049	105	59.6	NOT REPORTED
053	60	8.8	PLASMA, INDUCTIVELY COUPLED

TOTAL RANGE 50
 STANDARD DEVIATION 17.5
 MEAN: 65.8
 95 % CONFIDENCE INTERVAL OF MEAN 65.8 + OR - 10.6

TABLE 7.—STANDARD REFERENCE SAMPLE T-87 REPORT FOR BE

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
003	v	10	IGNORED ATOMIC ABSORPTION, DIRECT
004	v	10	IGNORED ATOMIC ABSORPTION, DIRECT
006	v	0	IGNORED ATOMIC ABSORPTION, FLAMELESS
010	v	6	IGNORED PLASMA, INDUCTIVELY COUPLED
012	v	10	IGNORED NOT REPORTED
013	v	1	IGNORED PLASMA, INDUCTIVELY COUPLED
015	v	0	IGNORED ATOMIC ABSORPTION, FLAMELESS
027	v	1	PLASMA, INDUCTIVELY COUPLED
042	v	1	PLASMA, INDUCTIVELY COUPLED
045	v	0	IGNORED PLASMA, INDUCTIVELY COUPLED
053	v	1	IGNORED PLASMA, INDUCTIVELY COUPLED
TOTAL RANGE	1	10	MEAN: 1.0
STANDARD DEVIATION	1	1	

TABLE 7.—STANDARD REFERENCE SAMPLE 1-H7 REPORT FOR CD

REPORTED CUEF VALUE	PCT. DEV. FROM MEAN	METHODS
0.02	1	IGNORED ATOMIC ABSORPTION, FLAMELESS
0.03	5	IGNORED PLASMA, INDUCTIVELY COUPLED
0.04	1	IGNORED PLASMA, INDUCTIVELY COUPLED
0.06	0	100.0 REJECT ATOMIC ABSORPTION, FLAMELESS
0.07	12	600.0 REJECT ATOMIC ABSORPTION, DIRECT
0.09	6	250.0 REJECT ATOMIC ABSORPTION, DIRECT
0.10	0	100.0 ATOMIC ABSORPTION, DIRECT
0.11	0	100.0 ATOMIC ABSORPTION, DIRECT
0.12	10	100.0 IGNORED NOT REPORTED
0.13	1	* * * IGNORED PLASMA, INDUCTIVELY COUPLED
0.14	0	100.0 ATOMIC ABSORPTION, DIRECT
0.15	0	ATOMIC ABSORPTION, FLAMELESS
0.19	0	100.0 NOT REPORTED
0.20	0	100.0 ATOMIC ABSORPTION, DIRECT
0.22	0	* * * IGNORED ATOMIC ABSORPTION, FLAMELESS
0.23	4	800.0 REJECT NOT REPORTED
0.25	1	* * * IGNORED ATOMIC ABSORPTION, DIRECT
0.26	2	350.0 ATOMIC ABSORPTION, CHELATION
0.27	1	125.0 ATOMIC ABSORPTION, FLAMELESS
0.30	0	100.0 NOT REPORTED
0.53	1	* * * IGNORED ATOMIC ABSORPTION, FLAMELESS
0.54	0	125.0 PLASMA, INDUCTIVELY COUPLED
0.55	1	IGNORED ATOMIC ABSORPTION, DIRECT
0.57	2	350.0 IGNORED ATOMIC ABSORPTION, FLAMELESS
0.59	10	* * * IGNORED ATOMIC ABSORPTION, DIRECT
0.40	0	* * * IGNORED ATOMIC ABSORPTION, DIRECT
0.41	0	100.0 ATOMIC ABSORPTION, FLAMELESS
0.42	5	* * * IGNORED PLASMA, INDUCTIVELY COUPLED
0.43	2	350.0 ATOMIC ABSORPTION, DIRECT
0.45	0	PLASMA, INDUCTIVELY COUPLED
0.53	1	* * * IGNORED ATOMIC ABSORPTION, FLAMELESS
0.54	10	150.0 REJECT ATOMIC ABSORPTION, FLAMELESS
0.57	0	100.0 ATOMIC ABSORPTION, FLAMELESS
0.59	0	100.0 ATOMIC ABSORPTION, FLAMELESS
0.65	0	100.0 ANODIC STRIPPING VOLTMETRY

TOTAL RANGE:
STANDARD DEVIATIONMEAN: 0.4
95 % CONFIDENCE INTERVAL OF MEAN 0.4 + OR - 0.4

TABLE 7.— STANDARD REFERENCE SAMPLE 1-87 REPORT FOR CO

REPORTED CURE VALUE	PC1. DEV. FROM MEAN	METHODS
0.02	5	*** IGNORED ATOMIC ABSORPTION, FLAMELESS
0.03	10	*** IGNORED PLASMA, INDUCTIVELY COUPLED
0.04	50	*** IGNORED ATOMIC ABSORPTION, DIRECT
0.06	2	*** IGNORED ATOMIC ABSORPTION, FLAMELESS
0.09	7	30.0 *** IGNORED ATOMIC ABSORPTION, DIRECT
0.10	2	*** IGNORED ATOMIC ABSORPTION, DIRECT
0.12	10	*** IGNORED ATOMIC ABSORPTION, DIRECT
0.13	1	*** IGNORED PLASMA, INDUCTIVELY COUPLED
0.14	1	*** IGNORED ATOMIC ABSORPTION, DIRECT
0.15	1	*** IGNORED ATOMIC ABSORPTION, FLAMELESS
0.19	0	100.0 *** IGNORED ATOMIC ABSORPTION, FLAMELESS NOT REPORTED
0.22	0	100.0 *** IGNORED ATOMIC ABSORPTION, FLAMELESS
0.25	3	*** IGNORED ATOMIC ABSORPTION, DIRECT
0.26	6	11.4 *** IGNORED ATOMIC ABSORPTION, DIRECT
0.27	5	7.1 *** IGNORED PLASMA, INDUCTIVELY COUPLED
0.30	0	*** IGNORED ATOMIC ABSORPTION, CHELATION PLASMA, INDUCTIVELY COUPLED
0.32	1	30.0 *** IGNORED ATOMIC ABSORPTION, FLAMELESS
0.33	7	122.9 *** IGNORED ATOMIC ABSORPTION, DIRECT
0.36	12	7.1 *** IGNORED ATOMIC ABSORPTION, DIRECT
0.37	5	*** IGNORED ATOMIC ABSORPTION, DIRECT
0.39	50	*** IGNORED ATOMIC ABSORPTION, DIRECT
0.40	1	*** IGNORED ATOMIC ABSORPTION, DIRECT
0.41	10	*** IGNORED ATOMIC ABSORPTION, FLAMELESS
0.42	10	*** IGNORED PLASMA, INDUCTIVELY COUPLED
0.43	7	30.0 *** IGNORED ATOMIC ABSORPTION, DIRECT
0.45	1	81.4 *** IGNORED PLASMA, INDUCTIVELY COUPLED
0.53	3	44.3 *** IGNORED ATOMIC ABSORPTION, FLAMELESS
0.57	0	100.0 *** IGNORED ATOMIC ABSORPTION, FLAMELESS
0.59	17	215.7 *** IGNORED ATOMIC ABSORPTION, FLAMELESS
TOTAL RANGE	0	10
STANDARD DEVIATION	5.0	17

MEAN: 5.4
 95 % CONFIDENCE INTERVAL OF MEAN 5.4 ± OR - 3.0

TABLE 7.—STANDARD REFERENCE SAMPLE I-87 REPORT FOR CR 101

REPORTED CODE	VALUE	PCI.	DEV. FROM MEAN	METHODS
002	<	5	***	IGNORED ATOMIC ABSORPTION, FLAMELESS
003	<	5	***	IGNORED PLASMA, INDUCTIVELY COUPLED
004	<	1.0	***	IGNORED PLASMA, INDUCTIVELY COUPLED
006	<	3	2.1	ATOMIC ABSORPTION, FLAMELESS
008	<	6	104.3	ATOMIC ABSORPTION, DIRECT
009	<	1	***	IGNORED ATOMIC ABSORPTION, DIRECT
010	<	5	***	IGNORED ATOMIC ABSORPTION, DIRECT
011	<	5	70.2	ATOMIC ABSORPTION, FLAMELESS
012	<	1.0	***	IGNORED NOT REPUTED
013	<	3	***	IGNORED PLASMA, INDUCTIVELY COUPLED
014	<	1	66.0	ATOMIC ABSORPTION, DIRECT
015	<	1	66.0	ATOMIC ABSORPTION, FLAMELESS
019	<	1	66.0	NOT REPUTED
022	<	0	100.0	ATOMIC ABSORPTION, FLAMELESS
025	<	1	***	IGNORED ATOMIC ABSORPTION, DIRECT
026	<	5	70.2	ATOMIC ABSORPTION, FLAMELESS
027	<	5	70.2	PLASMA, INDUCTIVELY COUPLED
030	<	0	100.0	NOT REPUTED
033	<	7	138.3	PLASMA, INDUCTIVELY COUPLED
034	<	0	100.0	ATOMIC ABSORPTION, FLAMELESS
035	<	1	***	IGNORED ATOMIC ABSORPTION, DIRECT
036	<	5	70.2	ATOMIC ABSORPTION, DIRECT
037	<	7.5	385.1	REJECT ATOMIC ABSORPTION, DIRECT
039	<	50	***	IGNORED ATOMIC ABSORPTION, DIRECT
040	<	1	***	IGNORED ATOMIC ABSORPTION, DIRECT
041	<	1	66.0	ATOMIC ABSORPTION, FLAMELESS
042	<	1.5	410.6	PLASMA, INDUCTIVELY COUPLED
043	<	2	***	IGNORED ATOMIC ABSORPTION, DIRECT
045	<	1	104.3	NOT REPUTED
049	<	6	***	IGNORED NOT REPUTED
051	<	1	***	IGNORED ATOMIC ABSORPTION, FLAMELESS
053	<	1	66.0	ATOMIC ABSORPTION, FLAMELESS
057	<	1	***	IGNORED ATOMIC ABSORPTION, FLAMELESS
059	<	3	***	IGNORED ATOMIC ABSORPTION, FLAMELESS

TOTAL RANGE:
STANDARD DEVIATIONMEAN: 2.9
95 % CONFIDENCE INTERVAL OF MEAN 2.9 + OR - 1.4

TABLE 7.— STANDARD REFERENCE SAMPLE T-67 REPORT FOR CU

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
002	<	5	*** IGNORED ATOMIC ABSORPTION, FLAMELESS
003	<	10	*** IGNORED PLASMA, INDUCTIVELY COUPLED
004	<	5	*** IGNORED PLASMA, INDUCTIVELY COUPLED
006	7	8.1	ATOMIC ABSORPTION, FLAMELESS
007	70	818.9	REJECT ATOMIC ABSORPTION, DIRECT
008	7	8.1	ATOMIC ABSORPTION, DIRECT
009	23	201.9	ATOMIC ABSORPTION, DIRECT
010	5	34.4	ATOMIC ABSORPTION, DIRECT
011	9	18.1	ATOMIC ABSORPTION, DIRECT
012	10	*** IGNORED NOT REPORTED	PLASMA, INDUCTIVELY COUPLED
013	4	47.5	ATOMIC ABSORPTION, DIRECT
014	6	5.0	ATOMIC ABSORPTION, FLAMELESS
015	4	47.5	ATOMIC ABSORPTION, FLAMELESS
018	5	34.4	ATOMIC ABSORPTION, FLAMELESS
019	5	34.4	NOT REPORTED
020	7	8.1	ATOMIC ABSORPTION, DIRECT
022	5	34.4	ATOMIC ABSORPTION, FLAMELESS
023	24	215.1	NOT REPORTED
025	4	47.5	ATOMIC ABSORPTION, DIRECT
026	9	18.1	ATOMIC ABSORPTION, DIRECT
027	4	47.5	ATOMIC ABSORPTION, DIRECT
030	0	100.0	NOT REPORTED
032	4	47.5	ATOMIC ABSORPTION, CHELATION
033	8	5.0	PLASMA, INDUCTIVELY COUPLED
034	4	47.5	ATOMIC ABSORPTION, FLAMELESS
035	4	47.5	ATOMIC ABSORPTION, DIRECT
036	11	44.4	ATOMIC ABSORPTION, DIRECT
037	7	8.1	ATOMIC ABSORPTION, DIRECT
039	<	*** IGNORED ATOMIC ABSORPTION, DIRECT	PLASMA, INDUCTIVELY COUPLED
040	5	34.4	ATOMIC ABSORPTION, DIRECT
041	8	5.0	ATOMIC ABSORPTION, FLAMELESS
042	<	*** IGNORED PLASMA, INDUCTIVELY COUPLED	ATOMIC ABSORPTION, DIRECT
043	6	21.2	ATOMIC ABSORPTION, DIRECT
045	9	18.1	ATOMIC ABSORPTION, DIRECT
049	4	47.5	NOT REPORTED
051	19	149.4	ATOMIC ABSORPTION, DIRECT
053	6	21.2	ATOMIC ABSORPTION, FLAMELESS
054	20	162.5	ATOMIC ABSORPTION, DIRECT
057	6	21.2	ATOMIC ABSORPTION, FLAMELESS
059	4	47.5	ATOMIC ABSORPTION, FLAMELESS
065	4	47.5	ANODIC STRIPPING VOLATAMMETRY
TOTAL RANGE	0	70	MEAN: 7.6
STANDARD DEVIATION	5.6	95 % CONFIDENCE INTERVAL OF MEAN	7.6 + OR - 2.0

TABLE 7.—STANDARD REFERENCE SAMPLE T-87 REPORT FOR FE

REPORTED CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	HIGH/LOW
002	< 20	***	IGNORED PLASMA, INDUCTIVELY COUPLED
003	< 30	***	IGNORED PLASMA, INDUCTIVELY COUPLED
004	80	177.6	PLASMA, INDUCTIVELY COUPLED
006	< 2	***	IGNORED ATOMIC ABSORPTION, FLAMELESS
008	11	61.8	ATOMIC ABSORPTION, DIRECT
010	< 1	***	IGNORED ATOMIC ABSORPTION, DIRECT
012	< 10	***	IGNORED NOT REPORTED
013	1	96.5	PLASMA, INDUCTIVELY COUPLED
014	4	86.1	ATOMIC ABSORPTION, DIRECT
015	40	38.9	COLORIMETRIC
019	30	4.2	NOT REPORTED
020	100	247.2	ATOMIC ABSORPTION, DIRECT
022	5	82.6	ATOMIC ABSORPTION, DIRECT
023	> 70	143.1	NOT REPORTED
025	< 50	***	IGNORED ATOMIC ABSORPTION, DIRECT
026	400	288.9	REJECT PLASMA, INDUCTIVELY COUPLED
027	5	82.6	PLASMA, INDUCTIVELY COUPLED
030	0	100.0	NOT REPORTED
032	12	58.3	ATOMIC ABSORPTION, FLAMELESS
035	< 10	***	IGNORED ATOMIC ABSORPTION, DIRECT
036	42	45.8	ATOMIC ABSORPTION, DIRECT
037	220	663.9	REJECT ATOMIC ABSORPTION, DIRECT
039	< 20	***	IGNORED ATOMIC ABSORPTION, DIRECT
041	30	4.2	ATOMIC ABSORPTION, DIRECT
042	6	79.2	PLASMA, INDUCTIVELY COUPLED
043	< 10	***	ATOMIC ABSORPTION, DIRECT
045	7	75.7	PLASMA, INDUCTIVELY COUPLED
049	13	54.9	COLORIMETRIC
051	40	38.9	COLORIMETRIC
053	10	65.3	PLASMA, INDUCTIVELY COUPLED
054	70	143.1	ATOMIC ABSORPTION, DIRECT
059	< 3	***	IGNORED ATOMIC ABSORPTION, FLAMELESS

TOTAL RANGE 0 TO 400 MEAN: 28.0
 STANDARD DEVIATION 30.0 95 % CONFIDENCE INTERVAL OF MEAN 28.0 ± 0.4 - 14.0

TABLE 7.— STANDARD REFERENCE SAMPLE T-67 REPORT FOR HG

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
003	< 0.1	*** 250.0	IGNORED ATOMIC ABSORPTION, COLD VAPOR
004	1.8	***	ATOMIC ABSORPTION, COLD VAPOR
006	< 0.2	***	IGNORED ATOMIC ABSORPTION, COLD VAPOR
U10	0.1	80.6	ATOMIC ABSORPTION, DIRECT
013	< 0.2	***	IGNORED ATOMIC ABSORPTION, COLD VAPOR
014	< 0.5	***	IGNORED ATOMIC ABSORPTION, DIRECT
015	0.1	60.6	ATOMIC ABSORPTION, DIRECT
U19	1.0	94.4	NOT REPORTED
023	10.0	REJECT	NOT REPORTED
025	0.3	41.7	ATOMIC ABSORPTION, DIRECT
027	0.2	61.1	OTHER
035	< 0.1	***	IGNORED ATOMIC ABSORPTION, DIRECT
037	< 1.0	***	IGNORED ATOMIC ABSORPTION, COLD VAPOR
043	< 0.5	***	IGNORED ATOMIC ABSORPTION, DIRECT
045	< 0.2	***	IGNORED NOT REPORTED
053	< 0.5	***	IGNORED NOT REPORTED
U57	0.1	60.6	ATOMIC ABSORPTION, COLD VAPOR
059	< 0.1	***	IGNORED ATOMIC ABSORPTION, COLD VAPOR
TOTAL RANGE	0.1	10.0	MEAN: 0.51
STANDARD DEVIATION	0.065	95 % CONFIDENCE INTRVL OF MEAN	0.51 + OR - 0.60

TABLE 7.--STANDARD REFERENCE SAMPLE T-67 REPORT FOR LI

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS	
			IGNORED	ATOMIC ABSORPTION, DIRECT
003	<	10	***	ATOMIC ABSORPTION, DIRECT
004	30	81.8	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
006	10	39.4	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
008	18	9.1	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
010	<	50	***	ATOMIC ABSORPTION, DIRECT
012	267	518.2	REJECT	NOT REPORTED
013	17	3.0	PLASMA, INDUCTIVELY COUPLED	PLASMA, INDUCTIVELY COUPLED
014	16	3.0	NOT REPORTED	NOT REPORTED
015	17	3.0	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
022	20	21.2	NOT REPORTED	NOT REPORTED
025	10	39.4	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
026	100	506.1	REJECT	EMISSION, FLAME
036	17	3.0	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
039	20	21.2	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
049	13	21.2	NOT REPORTED	NOT REPORTED
053	<	50	***	ATOMIC ABSORPTION, DIRECT
054	10	39.4	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT

TOTAL RANGE 10 10 267 MEAN: 16.5
 STANDARD DEVIATION 5.6 95 % CONFIDENCE INTRVL OF MEAN 16.5 + OR - 3.6

TABLE 7.-- STANDARD REFERENCE SAMPLE T-67 REPORT FOR MN

REPORTED CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS	
002	<	20	IGNORED PLASMA, INDUCTIVELY COUPLED	
003	<	10	IGNORED PLASMA, INDUCTIVELY COUPLED	
004	30	706.0	REJECT PLASMA, INDUCTIVELY COUPLED	
006	1	73.1	ATOMIC ABSORPTION, FLAMELESS	
008	4	7.5	ATOMIC ABSORPTION, DIRECT	
010	<	20	IGNORED ATOMIC ABSORPTION, DIRECT	
012	10	<*>	IGNORED NOT REPORTED	
013	1	73.1	PLASMA, INDUCTIVELY COUPLED	
014	1	73.1	ATOMIC ABSORPTION, DIRECT	
015	2	46.3	ATOMIC ABSORPTION, DIRECT	
019	10	168.7	NOT REPORTED	
020	2	46.3	ATOMIC ABSORPTION, DIRECT	
025	<	20	IGNORED ATOMIC ABSORPTION, DIRECT	
026	<	100	REJECT ATOMIC ABSORPTION, DIRECT	
027	<	1	73.1	PLASMA, INDUCTIVELY COUPLED
030	<	0	<*>	IGNORED NOT REPORTED
032	4	7.5	ATOMIC ABSORPTION, FLAMELESS	
035	<	10	IGNORED ATOMIC ABSORPTION, DIRECT	
056	4	7.5	ATOMIC ABSORPTION, DIRECT	
037	5	34.3	ATOMIC ABSORPTION, DIRECT	
039	<	30	IGNORED ATOMIC ABSORPTION, DIRECT	
040	1	73.1	ATOMIC ABSORPTION, DIRECT	
041	<	10	168.7	ATOMIC ABSORPTION, DIRECT
042	<	5	34.3	PLASMA, INDUCTIVELY COUPLED
043	<	2	IGNORED ATOMIC ABSORPTION, DIRECT	
045	1	73.1	PLASMA, INDUCTIVELY COUPLED	
049	9	141.8	NOT REPORTED	
051	<	4	<*>	IGNORED ATOMIC ABSORPTION, FLAMELESS
053	<	10	7.5	NOT REPORTED
059	2	46.3	ATOMIC ABSORPTION, FLAMELESS	

TOTAL RANGE 1 10 100 MEAN: 3.7
 STANDARD DEVIATION 3.1 95 % CONFIDENCE INTRVL OF MEAN 3.7 + OR - 1.5

TABLE 7.— STANDARD REFERENCE SAMPLE 1-87 REPORT FOR MO

CODE	REPORTED VALUE	PCT. DEV. FRM MEAN		METHODS
		PCT.	DEV.	
003	26	123.4	INDUCTIVELY COUPLED	PLASMA, INDUCTIVELY COUPLED
004	<	10	***	IGNORED PLASMA, INDUCTIVELY COUPLED
009	25	114.8	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
010	<	500	***	IGNORED PLASMA, INDUCTIVELY COUPLED
012	1162	865.9	REJECT	NOT REPUTED
014	5	57.0	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
020	4	65.6	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
025	5	57.0	ATOMIC ABSORPTION, DIRECT	ATOMIC ABSORPTION, DIRECT
026	<	10	***	IGNORED NOT REPUTED
027	9	22.7	PLASMA, INDUCTIVELY COUPLED	PLASMA, INDUCTIVELY COUPLED
032	8	31.2	ATOMIC ABSORPTION, FLAMELESS	ATOMIC ABSORPTION, FLAMELESS
033	5	57.0	PLASMA, INDUCTIVELY COUPLED	PLASMA, INDUCTIVELY COUPLED
042	29	149.2	PLASMA, INDUCTIVELY COUPLED	PLASMA, INDUCTIVELY COUPLED
045	6	48.4	PLASMA, INDUCTIVELY COUPLED	PLASMA, INDUCTIVELY COUPLED
049	6	48.4	NOT REPUTED	NOT REPUTED
053	<	5	***	IGNORED ATOMIC ABSORPTION, FLAMELESS

TOTAL RANGE 4 10 1162 MEAN: 11.6
 STANDARD DEVIATION 9.8 95 % CONFIDENCE INTVL OF MEAN 11.6 + OR - 6.6

TABLE 7.-- STANDARD REFERENCE SAMPLE 1-87 REPORT FOR NI

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
002	< 5	***	IGNORED ATOMIC ABSORPTION, FLAMELESS
003	< 10	***	IGNORED PLASMA, INDUCTIVELY COUPLED
004	25	194.1	IGNORED PLASMA, INDUCTIVELY COUPLED
006	3	64.7	ATOMIC ABSORPTION, FLAMELESS
008	7	17.6	ATOMIC ABSORPTION, DIRECT
009	25	194.1	ATOMIC ABSORPTION, DIRECT
010	< 2	***	IGNORED ATOMIC ABSORPTION, DIRECT
012	10	***	IGNORED NOT REPORTED
013	2	***	IGNORED PLASMA, INDUCTIVELY COUPLED
014	2	76.5	ATOMIC ABSORPTION, DIRECT
015	16	66.2	ATOMIC ABSORPTION, FLAMELESS
019	2	76.5	ATOMIC ABSORPTION, DIRECT
020	2	76.5	ATOMIC ABSORPTION, DIRECT
022	1	88.2	ATOMIC ABSORPTION, FLAMELESS
023	62	629.4	REJECT NOT REPORTED
025	< 3	***	IGNORED ATOMIC ABSORPTION, DIRECT
026	8	5.9	ATOMIC ABSORPTION, CHELATION
027	20	135.3	PLASMA, INDUCTIVELY COUPLED
030	0	100.0	NOT REPORTED
032	4	52.9	ATOMIC ABSORPTION, CHELATION
033	6	29.4	PLASMA, INDUCTIVELY COUPLED
034	< 5	***	IGNORED ATOMIC ABSORPTION, FLAMELESS
035	2	76.5	ATOMIC ABSORPTION, DIRECT
037	6	29.4	ATOMIC ABSORPTION, DIRECT
039	< 50	***	IGNORED ATOMIC ABSORPTION, DIRECT
040	< 1	***	IGNORED ATOMIC ABSORPTION, DIRECT
042	7	17.6	PLASMA, INDUCTIVELY COUPLED
043	5	41.2	ATOMIC ABSORPTION, DIRECT
045	3	64.7	PLASMA, INDUCTIVELY COUPLED
049	2	76.5	NOT REPORTED
051	17	100.0	ATOMIC ABSORPTION, DIRECT
053	< 2	***	IGNORED ATOMIC ABSORPTION, FLAMELESS
054	20	135.3	ATOMIC ABSORPTION, DIRECT
057	1	88.2	ATOMIC ABSORPTION, FLAMELESS
059	20	135.3	ATOMIC ABSORPTION, FLAMELESS

TOTAL RANGE 0 TO 100.0
 STANDARD DEVIATION 0.8.3
 MEAN: 8.5
 95 % CONFIDENCE INTERVAL OF MEAN 8.5 + OR - 3.5

TABLE 7.— STANDARD REFERENCE SAMPLE T-87 REPORT FOR PG

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
002	<	5	IGNORED ATOMIC ABSORPTION, FLAMELESS IGNORED PLASMA, INDUCTIVELY COUPLED
003	<	2	IGNORED PLASMA, INDUCTIVELY COUPLED
004	29	310.6	IGNORED ATOMIC ABSORPTION, FLAMELESS
006	<	2	IGNORED ATOMIC ABSORPTION, DIRECT
008	14	102.1	ATOMIC ABSORPTION, DIRECT
009	21	203.1	ATOMIC ABSORPTION, DIRECT
U10	0	***	IGNORED ATOMIC ABSORPTION, DIRECT
011	1	***	IGNORED ATOMIC ABSORPTION, DIRECT
013	10	***	IGNORED PLASMA, INDUCTIVELY COUPLED
014	1	***	IGNORED ATOMIC ABSORPTION, DIRECT
015	1	85.6	ATOMIC ABSORPTION, DIRECT
019	1	85.6	ATOMIC ABSORPTION, FLAMELESS
U22	0	100.0	NOT REPORTED
023	55	693.8	ATOMIC ABSORPTION, FLAMELESS
025	<	5	REJECT NOT REPORTED
026	<	10	IGNORED ATOMIC ABSORPTION, DIRECT
027	1	05.6	IGNORED NOT REPORTED
030	<	0	ATOMIC ABSORPTION, FLAMELESS
033	7	1.0	IGNORED NOT REPORTED
034	3	56.7	PLASMA, INDUCTIVELY COUPLED
035	<	1	ATOMIC ABSORPTION, FLAMELESS
037	7	***	IGNORED ATOMIC ABSORPTION, DIRECT
039	50	1.0	ATOMIC ABSORPTION, DIRECT
040	1	***	IGNORED ATOMIC ABSORPTION, DIRECT
041	1	***	IGNORED ATOMIC ABSORPTION, FLAMELESS
042	10	***	IGNORED PLASMA, INDUCTIVELY COUPLED
043	30	***	IGNORED ATOMIC ABSORPTION, DIRECT
045	3	***	IGNORED PLASMA, INDUCTIVELY COUPLED
049	2	71.1	NOT REPORTED
051	11	58.8	ATOMIC ABSORPTION, DIRECT
053	1	***	IGNORED ATOMIC ABSORPTION, FLAMELESS
057	0	100.0	ATOMIC ABSORPTION, FLAMELESS
065	0	100.0	ANODIC STRIPPING VOLAMMETRY

TOTAL RANGE 0 TO 100.0
 STANDARD DEVIATION 6.9
 MEAN: 6.9
 95 X CONFIDENCE INTRVL OF MEAN 6.9 + OR - 5.2

TABLE 7.-- STANDARD REFERENCE SAMPLE T-87 REPORT FOR SB

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN		METHODS
		***	0.0	
004	<	5	***	IGNORED ATOMIC ABSORPTION, HYDRIDE
006		1	0.0	ATOMIC ABSORPTION, HYDRIDE
010	<	2	***	IGNORED ATOMIC ABSORPTION, DIRECT
014	<	1	***	IGNORED ATOMIC ABSORPTION, DIRECT
015		1	0.0	IGNORED ATOMIC ABSORPTION, HYDRIDE
026	<	10	***	IGNORED ATOMIC ABSORPTION, DIRECT
027		50	900.0	REJECT PLASMA, INDUCTIVELY COUPLED
040	<	1	***	IGNORED ATOMIC ABSORPTION, DIRECT
053	<	2	***	IGNORED ATOMIC ABSORPTION, HYDRIDE
057		1	0.0	ATOMIC ABSORPTION, HYDRIDE
TOTAL RANGE		1	10	MEAN:
STANDARD DEVIATION		50	50	1.0

TABLE 7.— STANDARD REFERENCE SAMPLE T-87 REPORT FOR SE

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
003	<	5.0	IGNORED ATOMIC ABSORPTION, HYDRIDE
004	<	10.0	IGNORED ATOMIC ABSORPTION, HYDRIDE
006	<	2.0	ATOMIC ABSORPTION, HYDRIDE
009	<	10.0	IGNORED ATOMIC ABSORPTION, DIRECT
011	<	3.0	IGNORED ATOMIC ABSORPTION, FLAMELESS
015	0.5	58.3	ATOMIC ABSORPTION, HYDRIDE
025	1.0	16.7	ATOMIC ABSORPTION, DIRECT
027	0.5	58.3	ATOMIC ABSORPTION, FLAMELESS
033	10.0	733.3	REJECT PLASMA, INDUCTIVELY COUPLED
040	<	1.0	IGNORED ATOMIC ABSORPTION, DIRECT
042	<	5.0	IGNORED PLASMA, INDUCTIVELY COUPLED
045	<	3.0	IGNORED ATOMIC ABSORPTION, HYDRIDE
053	2.0	66.7	ATOMIC ABSORPTION, HYDRIDE

TOTAL RANGE 0.5 TO 10.0 MEAN: 1.20
 STANDARD DEVIATION 0.76 95 % CONFIDENCE INTRVL OF MEAN 1.20 + OR - 0.94

TABLE 7.-- STANDARD REFERENCE SAMPLE T-87 MEASURED FOR SR

CODE	REPORTED VALUE	PCT. FROM MEAN	DEV. FROM MEAN	METHODS
U03	760	2.3		PLASMA, INDUCTIVELY COUPLED
006	760	2.3		ATOMIC ABSORPTION, DIRECT
008	800	7.7		ATOMIC ABSORPTION, DIRECT
009	940	26.6		ATOMIC ABSORPTION, DIRECT
010	700	5.7		ATOMIC ABSORPTION, DIRECT
012	782	5.3		PLASMA, DIRECT CURRENT
013	760	2.3		PLASMA, INDUCTIVELY COUPLED
014	780	5.0		ATOMIC ABSORPTION, DIRECT
015	750	1.0		OTHER
022	690	7.1		ATOMIC ABSORPTION, DIRECT
025	640	13.6		ATOMIC ABSORPTION, DIRECT
027	735	1.0		PLASMA, INDUCTIVELY COUPLED
036	500	32.7		ATOMIC ABSORPTION, DIRECT
039	750	1.0		ATOMIC ABSORPTION, DIRECT
041	700	5.7		ATOMIC ABSORPTION, DIRECT
042	920	23.9		PLASMA, INDUCTIVELY COUPLED
045	710	4.4		PLASMA, INDUCTIVELY COUPLED
049	908	22.3		NOT REPORTED
053	750	1.0		PLASMA, INDUCTIVELY COUPLED
054	520	30.0		ATOMIC ABSORPTION, DIRECT
059	740	0.4		ATOMIC ABSORPTION, DIRECT

TOTAL RANGE 500 10 940 MEAN: 742.6
 STANDARD DEVIATION 107.6 95 % CONFIDENCE INTRVL OF MEAN 742.6 + OR - 49.0

TABLE 7.— STANDARD REFERENCE SAMPLE 1-87 REPORT FOR TL

CODE	REPORTED VALUE	PCT. FROM MEAN	DEV. METHODS
010	<	1	*** IGNORED PLASMA, INDUCTIVELY COUPLED
015	<	0	*** IGNORED OTHER
027	2	0.0	PLASMA, INDUCTIVELY COUPLED
040	<	3	*** IGNORED ATOMIC ABSORPTION, DIRECT
053	<	2	*** IGNORED ATOMIC ABSORPTION

TABLE 7.— STANDARD REFERENCE SAMPLE T-87 REPORT FOR ZN

CODE	REPURPOSED VALUE	PCT. DEV. FROM MEAN	METHODS
002	< 10	***	IGNORED PLASMA, INDUCTIVELY COUPLED
003	< 10	***	IGNORED PLASMA, INDUCTIVELY COUPLED
004	3.8	305.1	REJECT PLASMA, INDUCTIVELY COUPLED
006	7	25.4	ATOMIC ABSORPTION, DIRECT
007	8	14.7	ATOMIC ABSORPTION, DIRECT
008	1.5	59.9	ATOMIC ABSORPTION, DIRECT
009	1.2	27.9	NOT REPORTED
010	< 20	***	IGNORED ATOMIC ABSORPTION, DIRECT
011	1.4	49.3	ATOMIC ABSORPTION, DIRECT
012	< 10	***	IGNORED NOT REPORTED
013	5	46.7	PLASMA, INDUCTIVELY COUPLED
014	6	36.0	ATOMIC ABSORPTION, DIRECT
015	4	57.4	NOT REPORTED
018	7	25.4	ATOMIC ABSORPTION, FLAMELESS
019	6	14.7	NOT REPORTED
022	11	17.3	ATOMIC ABSORPTION, DIRECT
023	2.2	134.6	NOT REPORTED
025	6	36.0	ATOMIC ABSORPTION, DIRECT
026	9	4.0	ANODIC STRIPPING VOLAMMETRY, DIFFERENTIAL PULSE
027	7	25.4	PLASMA, INDUCTIVELY COUPLED
030	0	100.0	NOT REPORTED
032	5	46.7	ATOMIC ABSORPTION, DIRECT
033	7	25.4	PLASMA, INDUCTIVELY COUPLED
034	< 10	***	IGNORED ATOMIC ABSORPTION, DIRECT
035	1.5	59.9	ATOMIC ABSORPTION, DIRECT
036	1.4	49.3	ATOMIC ABSORPTION, DIRECT
037	8	14.7	ATOMIC ABSORPTION, DIRECT
039	< 20	***	IGNORED ATOMIC ABSORPTION, DIRECT
040	2	78.7	ATOMIC ABSORPTION, DIRECT
041	5	46.7	ATOMIC ABSORPTION, DIRECT
042	1.5	59.9	PLASMA, INDUCTIVELY COUPLED
043	1.2	27.9	ATOMIC ABSORPTION, DIRECT
045	1.6	70.6	PLASMA, INDUCTIVELY COUPLED
051	1.2	27.9	ATOMIC ABSORPTION, DIRECT
053	1.4	49.3	PLASMA, INDUCTIVELY COUPLED
054	12.0	179.4	REJECT ATOMIC ABSORPTION, DIRECT
057	6	36.0	ATOMIC ABSORPTION, DIRECT
TOTAL RANGE	0	10	120
STANDARD DEVIATION	4.9	9.4	MEANS: 9.4
		95 % CONFIDENCE INTERVAL OF MEAN	9.4 + OR - 1.9

TABLE 8.--STATISTICS BY METHOD FOR SAMPLE: T-87

DETERMINATION: ACID/CACUS			
METHOD	MEAN	STD DEV	N
TITRATION	396.6	16.7	12
***** OVER ALL *****	394.6	16.0	16
DETERMINATION: AG			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT	2.4	1.9	7
***** OVER ALL *****	1.9	1.6	10
DETERMINATION: AL			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT	13.6	13.3	4
PLASMA, INDUCTIVELY COUPLED	21.0	25.4	3
***** OVER ALL *****	30.1	34.5	12
DETERMINATION: AS			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT	2.0	0.0	3
ATOMIC ABSORPTION, HYDROGEN	2.2	0.4	5
***** OVER ALL *****	2.0	0.5	10
DETERMINATION: BA			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT	66.7	28.9	3
PLASMA, INDUCTIVELY COUPLED	62.3	6.2	7
***** OVER ALL *****	65.6	17.5	13
DETERMINATION: BE			
METHOD	MEAN	STD DEV	N
***** INSUFFICIENT DATA *****			

TABLE 6.--STATISTICS BY METHOD FOR SAMPLE: T-67

DETERMINATION: CD			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT	0.7	1.0	6
ATOMIC ABSORPTION, FLAMELESS	0.2	0.4	6
***** OVER ALL *****	0.4	0.6	18
DETERMINATION: CO			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT	7.4	2.7	5
ATOMIC ABSORPTION, FLAMELESS	5.0	0.1	4
PLASMA, INDUCTIVELY COUPLED	4.3	3.1	3
***** OVER ALL *****	5.4	5.0	13
DETERMINATION: CR TOT			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT	4.0	2.6	3
ATOMIC ABSORPTION, FLAMELESS	1.6	1.8	7
***** OVER ALL *****	2.9	2.5	16
DETERMINATION: CU			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT	9.6	6.1	15
ATOMIC ABSORPTION, FLAMELESS	5.3	1.4	10
PLASMA, INDUCTIVELY COUPLED	7.0	2.6	3
***** OVER ALL *****	7.6	5.6	34
DETERMINATION: FT			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT	37.4	36.3	7
COLORIMETRIC	31.0	15.6	3
PLASMA, INDUCTIVELY COUPLED	16.2	30.4	6
***** OVER ALL *****	28.6	30.0	20

TABLE 8.--STATISTICS BY METHOD FOR SAMPLE: T-87

DETERMINATION: HG			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT ***** OVER ALL *****	0.17 0.51	0.12 0.65	3 7
 DETERMINATION: LI			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT ***** OVER ALL *****	16.5 16.5	6.8 5.6	8 12
 DETERMINATION: MN			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT ATOMIC ABSORPTION, FLAMELESS PLASMA, INDUCTIVELY COUPLED ***** OVER ALL *****	3.6 2.3 2.0 3.7	3.0 1.5 2.0 3.1	6 3 4 10
 DETERMINATION: MU			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT PLASMA, INDUCTIVELY COUPLED ***** OVER ALL *****	9.8 15.0 11.6	10.2 11.6 9.8	4 5 11
 DETERMINATION: NI			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT ATOMIC ABSORPTION, FLAMELESS PLASMA, INDUCTIVELY COUPLED ***** OVER ALL *****	9.6 8.2 12.2 8.5	8.8 9.1 9.7 8.3	9 5 5 24

TABLE 8.--STATISTICS BY METHOD FOR SAMPLE: T-87

DETERMINATION: PH	METHOD	MEAN	STD DEV	N
	ATOMIC ABSORPTION, DIRECT	13.3	5.9	4
	ATOMIC ABSORPTION, FLAMELESS	1.0	1.2	5
	***** OVER ALL *****	6.9	8.9	14
DETERMINATION: SB	METHOD	MEAN	STD DEV	N
	ATOMIC ABSORPTION, HYDRIDE	1.0	0.0	3
	***** OVER ALL *****	1.0	0.0	3
DETERMINATION: SE	METHOD	MEAN	STD DEV	N
	ATOMIC ABSORPTION, HYDRIDE	1.50	0.87	3
	***** OVER ALL *****	1.20	0.76	5
DETERMINATION: SR	METHOD	MEAN	STD DEV	N
	ATOMIC ABSORPTION, DIRECT	710.9	125.1	11
	PLASMA, INDUCTIVELY COUPLED	762.1	73.5	7
	***** OVER ALL *****	742.6	107.6	21
DETERMINATION: TL	METHOD	MEAN	STD DEV	N
	***** INSUFFICIENT DATA *****			
DETERMINATION: ZN	METHOD	MEAN	STD DEV	N
	ATOMIC ABSORPTION, DIRECT	9.1	4.1	16
	PLASMA, INDUCTIVELY COUPLED	10.7	4.8	6
	***** OVER ALL *****	9.4	4.9	29

TABLE 9.—STANDARD REFERENCE SAMPLE P-5 REPORT FOR CA

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
001	0.27	1.7	ATOMIC ABSORPTION, DIRECT
002	0.51	12.9	ATOMIC ABSORPTION, DIRECT
003	0.25	9.0	PLASMA, INDUCTIVELY COUPLED
004	< 0.01	***	IGNORED PLASMA, INDUCTIVELY COUPLED
006	0.25	9.0	ATOMIC ABSORPTION, DIRECT
007	0.34	23.6	TITRATION
008	0.30	9.2	ATOMIC ABSORPTION, DIRECT
009	0.10	63.6	REJECT ATOMIC ABSORPTION, DIRECT
010	0.29	5.6	ATOMIC ABSORPTION, DIRECT
011	0.40	45.7	REJECT TITRATION
012	0.26	5.3	PLASMA, DIRECT CURRENT
013	0.26	5.3	PLASMA, INDUCTIVELY COUPLED
014	< 0.50	***	IGNORED ATOMIC ABSORPTION, DIRECT
015	0.26	5.3	ATOMIC ABSORPTION, DIRECT
017	< 0.01	***	IGNORED TITRATION
018	0.26	5.3	ATOMIC ABSORPTION, DIRECT
019	0.28	2.0	NOT REPUTED
020	< 0.01	***	IGNORED ATOMIC ABSORPTION, DIRECT
022	0.24	12.6	ATOMIC ABSORPTION, DIRECT
TOTAL RANGE	0.10	0.40	MEANS 0.275
STANDARD DEVIATION	0.026	95 % CONFIDENCE INTERVAL OF MEAN	0.275 + OR - 0.017

TABLE 9.— STANDARD REFERENCE SAMPLE P-5 REPORTED FOR CL

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
001	0.08	67.7	ION CHROMATOGRAPHY
002	0.10	59.6	COLORIMETRIC
003	0.10	***	IGNORED NOT REPORTED
004	0.30	21.1	TITRATION
006	0.08	67.7	ION CHROMATOGRAPHY
007	3.40	272.7	TITRATION
008	0.45	61.7	TITRATION
009	0.79	218.9	ION SPECIFIC ELECTRODE
010	0.10	59.6	ION CHROMATOGRAPHY
011	0.85	243.2	TITRATION
012	0.26	5.0	COLORIMETRIC
013	0.20	***	IGNORED COLORIMETRIC
014	5.00	***	IGNORED TITRATION
015	0.11	55.6	TITRATION
017	0.01	***	IGNORED TITRATION
018	0.00	100.0	TITRATION
019	0.01	96.0	NOT REPORTED
020	2.20	788.2	TITRATION
022	0.09	63.7	PLASMA, INDUCTIVELY COUPLED

TOTAL RANGE 0.00 TO 3.40 MEAN: 0.248
 STANDARD DEVIATION 0.283 95 % CONFIDENCE INTVL OF MEAN 0.248 ± OR - 0.171

TABLE 9.—STANDARD REFERENCE SAMPLE P-5 REPORT FOR F

CODE	REPURIFIED VALUE	PCT. DEV. FROM MEAN	METHODS
001	0.03	86.5	IUN SPECIFIC ELECTRODE
002	< 0.05	***	IGNORED IUN SPECIFIC ELECTRODE
003	< 0.02	***	IGNORED IUN SPECIFIC ELECTRODE
004	0.61	175.5	IUN SPECIFIC ELECTRODE
006	0.01	95.5	IUN CHROMATOGRAPHY
009	0.24	6.4	IUN SPECIFIC ELECTRODE
010	< 0.10	***	IGNORED IUN SPECIFIC ELECTRODE
011	0.04	81.9	IUN SPECIFIC ELECTRODE
012	0.60	171.0	NOT REPURIFIED
015	0.02	91.0	IUN SPECIFIC ELECTRODE
017	< 0.01	***	IGNORED IUN SPECIFIC ELECTRODE

TOTAL RANGE 0.01 STANDARD DEVIATION 0.010 MEAN: 0.221
 STANDARD DEVIATION UN 0.274 95 % CONFIDENCE INTERVAL OF MEAN 0.221 + OR - 0.253

TABLE 9.—STANDARD REFERENCE SAMPLE P-5 REPORT FOR K

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
001	0.04	0.0	ATOMIC ABSORPTION, DIRECT
002	0.04	0.0	ATOMIC ABSORPTION, DIRECT
003	0.50	***	IGNORED PLASMA, INDUCTIVELY COUPLED
004	<	0.10	IGNORED ATOMIC ABSORPTION, DIRECT
006	0.03	25.0	ATOMIC ABSORPTION, DIRECT
007	0.00	100.0	EMISSION, FLAME
008	0.12	200.0	ATOMIC ABSORPTION, DIRECT
010	<	0.10	IGNORED ATOMIC ABSORPTION, DIRECT
011	0.00	100.0	EMISSION, FLAME
012	<	0.05	IGNORED EMISSION, FLAME
013	0.07	75.0	PLASMA, INDUCTIVELY COUPLED
014	<	0.10	IGNORED NOT REPORTED
015	<	0.01	EMISSION, FLAME
017	<	0.01	IGNORED EMISSION, FLAME
018	0.05	25.0	EMISSION, FLAME
019	0.05	25.0	NOT REPORTED
020	0.31	675.0	REJECT EMISSION, FLAME
022	0.03	25.0	ATOMIC ABSORPTION, DIRECT
TOTAL RANGE	0.00	10	MEAN: 0.040
STANDARD DEVIATION	0.034	0.31	95 % CONFIDENCE INTERVAL OF MEAN 0.040 + OR - 0.023

TABLE 9.—STANDARD REFERENCE SAMPLE P-5 REPORT FOR Mg

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
001	0.02	27.3	ATOMIC ABSORPTION, DIRECT
002	0.02	27.3	ATOMIC ABSORPTION, DIRECT
003	0.05	61.6	PLASMA, INDUCTIVELY COUPLED
004	0.01	***	PLASMA, INDUCTIVELY COUPLED
006	0.02	27.3	ATOMIC ABSORPTION, DIRECT
007	0.05	61.6	LITRAILION
008	0.03	9.1	ATOMIC ABSORPTION, DIRECT
010	0.02	27.3	ATOMIC ABSORPTION, DIRECT
011	0.24	172.6	REJECT LITRAILION
012	0.01	***	IGNORED PLASMA, DIRECT CURRENT
013	0.02	27.3	PLASMA, INDUCTIVELY COUPLED
014	0.10	***	IGNORED ATOMIC ABSORPTION, DIRECT
015	0.02	27.3	ATOMIC ABSORPTION, DIRECT
017	0.01	***	IGNORED LITRAILION
018	0.02	27.3	ATOMIC ABSORPTION, DIRECT
019	0.04	45.5	NOT REPUTED
020	0.01	***	IGNORED ATOMIC ABSORPTION, DIRECT
022	0.02	27.3	ATOMIC ABSORPTION, DIRECT

TOTAL RANGE 0.02 TO 0.24
 STANDARD DEVIATION 0.012 MEAN: 0.027
 95 % CONFIDENCE INTERVAL OF MEAN 0.027 + OR - 0.008

TABLE 9.—STANDARD REFERENCE SAMPLE P-5 REPORT FOR NA

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
001	0.13	18.0	ATOMIC ABSORPTION, DIRECT
002	0.13	18.0	ATOMIC ABSORPTION, DIRECT
003	0.20	26.1	PLASMA, INDUCTIVELY COUPLED
004	< 0.01	***	IGNITED PLASMA, INDUCTIVELY COUPLED
006	0.13	18.0	ATOMIC ABSORPTION, DIRECT
007	9.20	701.8	EMISSION, FLAME
008	0.35	120.7	ATOMIC ABSORPTION, DIRECT
010	0.15	5.4	ATOMIC ABSORPTION, DIRECT
011	0.00	100.0	EMISSION, FLAME
012	0.12	24.3	EMISSION, FLAME
013	0.14	11.7	PLASMA, INDUCTIVELY COUPLED
014	< 0.10	***	IGNITED BUT NOT REPORTED
015	0.24	51.4	EMISSION, FLAME
017	< 0.04	***	IGNITED EMISSION, FLAME
018	0.10	36.9	ATOMIC ABSORPTION, DIRECT
019	0.16	0.9	NOT REPORTED
020	0.23	45.0	EMISSION, FLAME
022	0.14	11.7	ATOMIC ABSORPTION, DIRECT

TOTAL RANGE
STANDARD DEVIATIONMEAN:
0.159

95 % CONFIDENCE INTERVAL OF MEAN

0.159 + UR -

0.046

TABLE 9.—STANDARD REFERENCE SAMPLE P-5 REPORT FOR NH₃-N

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
001	0.01	0.0	COLORIMETRIC
002	0.02	***	IGNORED COLORIMETRIC
003	0.05	***	IGNORED ION SPECIFIC ELECTRODE
004	0.32	100.3	REJECT TITRATION
006	0.01	0.0	COLORIMETRIC
007	0.04	300.0	REJECT NOT REPORTED
009	2.22	102.2	REJECT COLORIMETRIC
010	0.02	***	IGNORED COLORIMETRIC
011	0.05	***	IGNORED COLORIMETRIC
012	0.05	***	IGNORED TITRATION
013	0.01	0.0	COLORIMETRIC
014	0.10	***	IGNORED COLORIMETRIC
015	0.01	***	IGNORED COLORIMETRIC
020	0.02	***	IGNORED NOT REPORTED
022	0.00	100.0	REJECT COLORIMETRIC

TOTAL RANGE 0.00 TO 2.22
 STANDARD DEVIATION 0.000 95 % CONFIDENCE INTRVL OF MEAN 0.010 + OR - 0.000

TABLE 9.--STANDARD REFERENCE SAMPLE P-5 REPORT FOR NBS-N

CODE	REPORTED VALUE	PCI.	DEV. FROM MEAN	METHODS	
				ION CHROMATOGRAPHY	NOT REPUTED
001	0.05	54.5	9.1		
003	0.10			COLORIMETRIC	
004	0.04	63.6		COLORIMETRIC	
006	0.07	36.4		COLORIMETRIC	
007	0.30	172.7		OTHER	
008	0.20	81.8		COLORIMETRIC	
009	1.98	700.0		COLORIMETRIC	
010	<	0.10	***	IGNORED ION CHROMATOGRAPHY	
011	<	0.50	***	IGNORED COLORIMETRIC	
012	<	0.06	***	IGNORED COLORIMETRIC	
014	<	0.10	***	IGNORED COLORIMETRIC	
015	<	0.10	***	IGNORED COLORIMETRIC	
017	<	0.02	***	IGNORED COLORIMETRIC	
018	0.05	54.5		COLORIMETRIC	
020	<	0.44	***	IGNORED COLORIMETRIC	
022	0.07	36.4		PLASMA, INDUCTIVELY COUPLED	

TOTAL RANGE 0.04 70 1.98 MEAN: 0.110
 STANDARD DEVIATION 0.092 95 % CONFIDENCE INTERVAL OF MEAN 0.110 + OR - 0.077

TABLE 9.— STANDARD REFERENCE SAMPLE P-5 REPORT FOR PH

CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
001	8.11	38.1	REJECT
002	5.95	1.3	ELECTROMETRIC
003	6.00	2.2	ELECTROMETRIC
004	6.15	4.7	NOT REPORTED
006	5.95	1.3	ELECTROMETRIC
007	6.10	3.9	ELECTROMETRIC
008	5.00	14.6	ELECTROMETRIC
009	5.40	8.0	NOT REPORTED
010	6.10	3.9	ELECTROMETRIC
011	6.35	8.1	NOT REPORTED
012	5.80	1.2	ELECTROMETRIC
013	6.40	9.0	ELECTROMETRIC
014	5.92	0.8	ELECTROMETRIC
015	5.50	6.3	ELECTROMETRIC
017	5.87	0.0	ELECTROMETRIC
018	5.46	7.0	ELECTROMETRIC
019	5.85	0.4	NOT REPORTED
020	6.05	3.0	ELECTROMETRIC
022	5.84	0.5	ELECTROMETRIC

TOTAL RANGE 5.00 MEAN: 5.872
 STANDARD DEVIATION 0.347 95 % CONFIDENCE INTERVAL OF MEAN 5.872 + OR - 0.173

TABLE 9.—STANDARD REFERENCE SAMPLE P-5 REPORT FOR SU4

CODE	REPORTED VALUE	PCI. FROM MEAN	DEV. FROM MEAN	METHODS
001	0.40	1.4	1.4	ION CHROMATOGRAPHY
002	0.30	26.0	26.0	COLUMNETRIC
003	0.35	13.7	13.7	PLASMA, INDUCTIVELY COUPLED
004	< 0.10	***	***	IGNORED
006	0.33	18.6	18.6	ION CHROMATOGRAPHY
008	1.55	282.2	REJECT	GRAVIMETRIC
010	< 0.30	***	***	IGNORED
011	5.29	711.2	REJECT	GRAVIMETRIC
012	< 0.05	***	***	IGNORED
013	0.37	8.8	8.8	PLASMA, INDUCTIVELY COUPLED
014	< 5.00	***	***	IGNORED ATOMIC ABSORPTION, DIRECT
015	0.50	23.3	23.3	GRAVIMETRIC
017	0.42	3.6	3.6	ION CHROMATOGRAPHY
019	0.62	52.9	52.9	NOT REPORTED
022	0.36	11.2	11.2	PLASMA, INDUCTIVELY COUPLED

TOTAL RANGE 0.30
 STANDARD DEVIATION 0.099 10 3.29 95 % CONFIDENCE INTERVAL OF MEAN MEAN: 0.406
 0.406 + OR - 0.076

TABLE 9.—STANDARD REFERENCE SAMPLE P-5 REPORT FOR SP. COND.

REPORTED CODE	REPORTED VALUE	PCT. DEV. FROM MEAN	METHODS
001	3.50	22.3	DIRECT READING INSTRUMENT
002	3.44	23.6	DIRECT READING INSTRUMENT
003	4.30	4.5	NOT REQUIRED
004	5.97	32.5	DIRECT READING INSTRUMENT
006	3.80	15.6	DIRECT READING INSTRUMENT
007	8.00	77.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER
008	7.00	55.4	DIRECT READING INSTRUMENT
009	10.00	453.9	REJECT NOT REQUIRED
010	2.00	55.6	DIRECT READING INSTRUMENT
011	6.00	77.6	NOT REQUIRED
012	3.47	23.0	DIRECT READING INSTRUMENT
014	6.00	33.2	DIRECT READING INSTRUMENT
015	4.00	11.2	NOT REQUIRED
017	2.90	35.6	DIRECT READING INSTRUMENT
018	3.00	33.4	DIRECT READING INSTRUMENT
019	4.00	11.2	NOT REQUIRED
020	5.70	17.9	DIRECT READING INSTRUMENT
022	5.50	22.3	DIRECT READING INSTRUMENT

TOTAL RANGE 2.00 10 70.00 95 % CONFIDENCE INTERVAL DF MEAN 4.505 + OR - 0.926
 STANDARD DEVIATION 1.804 MEAN: 4.505

TABLE 10.--STATISTICS BY METHOD FOR SAMPLE: P-5

DETERMINATION: CA			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT ***** OVER ALL *****	0.272	0.025	8
	0.275	0.028	13
DETERMINATION: CL			
METHOD	MEAN	STD DEV	N
ION CHROMATOGRAPHY TITRATION ***** OVER ALL *****	0.087	0.012	3
	0.342	0.333	5
	0.248	0.283	13
DETERMINATION: F			
METHOD	MEAN	STD DEV	N
ION SPECIFIC ELECTRODE ***** OVER ALL *****	0.188	0.253	5
	0.221	0.274	7
DETERMINATION: K			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT EMISSION, FLAME ***** OVER ALL *****	0.052	0.038	5
	0.015	0.024	4
	0.040	0.034	11
DETERMINATION: Mg			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT EMISSION, FLAME ***** OVER ALL *****	0.021	0.004	6
	0.027	0.012	12
DETERMINATION: Na			
METHOD	MEAN	STD DEV	N
ATOMIC ABSORPTION, DIRECT EMISSION, FLAME ***** OVER ALL *****	0.161	0.005	7
	0.147	0.112	4
	0.159	0.080	14

TABLE 10.—STATISTICS BY METHOD FOR SAMPLE: P-5

DETERMINATION: NH₃-N

METHOD	MEAN	STD DEV	N
COLORIMETRIC	0.010	0.000	3
***** OVER ALL *****	0.010	0.000	3

DETERMINATION: NO₂-N

METHOD	MEAN	STD DEV	N
COLORIMETRIC	0.090	0.074	4
***** OVER ALL *****	0.110	0.092	6

DETERMINATION: PH

METHOD	MEAN	STD DEV	N
ELECTROMETRIC	5.045	0.356	13
***** OVER ALL *****	5.072	0.347	10

DETERMINATION: SO₄

METHOD	MEAN	STD DEV	N
ION CHROMATOGRAPHY	0.383	0.047	3
PLASMA, INDUCTIVELY COUPLED	0.360	0.010	3
***** OVER ALL *****	0.406	0.099	9

DETERMINATION: SP. COND.

METHOD	MEAN	STD DEV	N
DIRECT READING INSTRUMENT	4.023	1.467	12
***** OVER ALL *****	4.505	1.804	17

Table 11.—Comparison of domestic and international laboratory analyses obtained on Standard Reference Water Sample M-86 (Major)

Constituent	Mean Concentration (mg/L) ^{1/}		Standard Deviation		Based on Analysis by Laboratories	
	Domestic	Int'l	Domestic	Int'l	Domestic	Int'l
Alk (CaCO ₃)	151	152	5.4	8.5	69	38
B	.240	.224	.075	.052	27	14
Br	.291	.544	.298	.285	11	5
Ca	70.6	70.5	6.8	4.0	70	44
Cl	44.8	44.4	1.7	3.1	69	43
DSRD 180°C	581	586	17	33	59	24
F	2.0	1.9	.21	.24	62	25
K	4.72	4.59	.59	.37	66	44
Mg	28.0	27.6	1.5	1.8	66	43
Na	77.0	77.0	3.3	3.8	69	44
NO ₂ -N	.01	.04	.004	.053	42	24
NO ₃ -N	3.98	3.86	.39	.38	68	34
P, Total	.50	.48	.066	.070	59	34
pH, units	8.13	7.86	.20	.18	84	46
SiO ₂	12.6	11.9	2.37	4.00	52	39
SO ₄ ²⁻	222	220	12.7	15.9	68	45
Sp. Cond.	859	860	48	53	71	42
Sr	753	765	97	98	20	16
V	18.1	10.8	12.1	1.1	19	9

^{1/} Except Sp. Cond. $\mu\text{S}/\text{cm}$.

Table 12.—Comparison of domestic and international laboratory analyses obtained on Standard Reference Water Sample T-87 (Trace)

Constituent	Mean Concentration ($\mu\text{g/L}$)		Standard Deviation		Based on Analysis by Laboratories	
	Domestic	Int'l	Domestic	Int'l	Domestic	Int'l
Acid (CaCO_3)	404	395	17	16	15	16
Ag	2.5	1.9	3.1	1.8	14	10
Al	78.8	30.5	114	34.1	12	12
As	4.7	2.0	4.6	.5	37	10
Ba	80.3	65.8	51.3	17.5	35	13
Be	0.5	1.0	0.16	0	4	2
Cd	.96	.4	1.1	.8	25	18
Co	5.0	5.4	5.1	5.0	7	13
Cr Tot	6.5	2.9	6.8	2.5	31	16
Cu	8.9	7.6	4.7	5.6	45	34
Fe	21.7	28.8	20.3	30.0	30	20
Hg	.28	.51	.12	.65	27	7
Li	25.8	16.5	7.6	5.6	19	12
Mn	5.4	3.7	5.4	3.1	24	18
Mo	7.5	11.6	1.4	9.8	11	11
Ni	8.9	8.5	8.3	8.3	24	24
Pb	4.3	6.9	3.9	8.9	28	14
Se	3.7	1.2	3.5	.8	29	5
Sr	753	743	120	108	23	21
Zn	11.5	9.4	6.1	4.9	44	29

Table 13.—Comparison of domestic and international laboratory analyses obtained on Standard Reference Water Sample P-5 (Precipitation-snowmelt)

Constituent	Mean Concentration (mg/L)		Standard Deviation		Based on Analysis by Laboratories	
	Domestic	Int'l	Domestic	Int'l	Domestic	Int'l
Ca	0.290	0.275	0.079	0.028	44	13
Cl	.324	.248	.324	.283	31	13
F	.027	.221	.025	.274	23	7
K	.053	.040	.051	.034	35	11
Mg	.032	.027	.034	.012	37	12
Na	.145	.159	.060	.080	41	14
NH ₃ -N	.022	.010	.023	.0	28	3
NO ₃ -N	.084	.110	.017	.092	42	8
pH units	5.83	5.87	.63	.35	50	18
SO ₄	.556	.406	.430	.099	27	9
Sp. Cond. $\mu\text{S}/\text{cm}$	4.28	4.50	1.40	1.80	45	17